



Development and Application of the Frequency-Scanning-Method for Filtered Rayleigh Scattering (FSM-FRS)

G. Stockhausen, U. Doll, C. Willert

DLR Cologne, Institute of Propulsion Technology


Engine Measurement Group (AT-OTM)



Outline

1. Filtered Rayleigh Scattering (FRS)
2. Frequency-Scanning-Method of FRS (FSM-FRS)
3. Applications of FSM-FRS
4. Summary and conclusions for FSM-FRS

1. Filtered Rayleigh Scattering (FRS) the principle





**on-chip integration produces one intensity value
loss of information problem...!**

**temperature, pressure, density,
Doppler frequency / velocity**

1. Filtered Rayleigh Scattering (FRS) suppressing laser reflex light


molecular absorption filter
(iodine vapor)



Enables measurement near walls or other objects!

1. Typical FRS setup


- wlm: wavelength meter
- smf: single mode fibre
- lso: light sheet optics
- roi: region of interest
- bd: beam dump
- fl1: focusing lens 1, variable
- fl2: focusing lens 2, 50mm, $f=1.3$
- ic: iodine cell
- bpf: bandpass filter, FWHM=1nm
- fl3: focusing lens 3, 100mm, $f=2$



Continuous wave laser system @ 532nm

Narrow linewidth (FWHM < 5MHz), tuning range 60GHz

2. Frequency scanning method of FRS



Trade-off:

Multi-Parameter +
accuracy

vs.


Measurement time
(mean values)

dependency of the FSM-FRS signal intensity (convolution of Rayleigh and iodine spectra) :

- **frequency shift (velocity component via Doppler effect)**
- **profile shape (width correlates with temperature, shape transforms with pressure)**
- **amplitude (number of scatter particles, density)**




2. Frequency scanning method of FRS







measurement of FRS spectrum for every camera pixel (ca. 128x128 up to 512x512)


3. Applications of FSM-FRS: A) Free jet experiment




A) Free jet experiment




B) Ranque-Hilsch vortex tube




set-up of the vortex tube



seeded flow and radial laser beam viewed from hot




seeded flow and axial laser sheet viewed side-on



comparison of FRS and L2F results

B) Ranque-Hilsch vortex tube





FRS results

CFD results

C) Atmospheric combustion




laboratory methane-air flame




$$T_{\text{FRS}}=1733 \text{ K} \leftrightarrow T_{\text{CARS}}=1765 \text{ K}$$

D) Quantification of cooling films




effusion cooled wall inside a
single sector combustor
(5 bar, preheated air @ 450K)




E) Determination of turbine entrance profiles

Image transfer with 4.5m flexible endoscope




Laser light transfer with 11m photonic crystal fiber



Traversing the laser beam allows reconstruction of complete measurement plane (interpolated)



FRS signal observed from endoscope

E) Determination of turbine entrance profiles



Successful test of endoscopic hardware and validation with PIV and wall temperature probes!




F) Development of a FRS point sensor



sensor dimensions: outer diameter=12mm, length=200mm

moving the sensor head allows to obtain multiple point measurements along the line of sight.

F) Development of a FRS point sensor



Performance: temperature +/- 1%, pressure 1-2% systematic deviation because of increasing defocussing with increasing pressure

4. Summary and conclusions for FSM-FRS


- further developed optical measurement technique for the determination of temperature, pressure and Doppler frequency shifts (stationary but planar)
- principle: spectral **f**iltering of **R**ayleigh-**s**cattered light through a molecular absorption filter (iodine vapour)
- the **f**requency **s**canning **m**ethod yields the convolution of the Rayleigh spectrum with an iodine absorption line
- suitable for measurements near walls or objects because laser reflexes are completely suppressed
- technique is compatible with endoscopic hardware for access in capsulated environments (usage of flexible fiber bundles for laser light & image transport)
- vortex core measurements are possible since no seeding particles are required (light scattered of flow molecules)
- difficulties: gas composition has to be known, pressure measurements are challenging



**Thank you
for your
attention!**



2. Filtered Rayleigh Scattering (FRS) determination of velocity



$$\Delta v = v - v_0 = \frac{v_0}{c} (\vec{o} - \vec{l}) \cdot \vec{v}$$

$$\text{Bsp.: } \vec{v}_z = 0, \vec{v}_y = 0$$

$$\Delta v = -105 \text{ MHz} \rightarrow \vec{v}_x \cong 65 \text{ m/s}$$

variations of observing direction \vec{o} and/or laser direction \vec{l} enables measurement of all 3 velocity components