Comparison of classification models for spectral data of laser-induced fluorescence

Poster 1.12

M. Kraus (1), L. Fellner (1), F. Gebert (1), K. Grünwald (1), C. Pargmann (1), A. Walter (1), F. Duschek (1)
(1) German Aerospace Center, Institute of Technical Physics, Langer Grund, Lampoldshausen,
74239 Hardthausen, Germany.

Deutsches Zentrum
für Luft- und Raumfahrt
German Aerospace Center

Preprocess

- Background correction
- Despiking, Smoothing, Rescaling
- Extraction of relevant ranges

Experimental Data

- 7 bacteria samples
- 3500 single spectra of 2 wavelengths
- Combination of corresponding spectra

Training

- Randomized samples for training and test
- Resampling: bootstrap and cross validation

- 2625 spectra for training, 875 for test
- Combination of several tuning parameters to find models with best performance

Simultaneous emission of two laser pulses with wavelengths of 266 and 355 nm. Shifted by optic delay line.
Distance to sample (indoor): 3.5 m. PMT-channels: 32. [2]

Averaged spectra for each bacteria with boundaries of 2nd and 3rd quartile

Setup #2 test results

Prediction	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
Reference	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
B.brevis	125	0	0	0	0	0	0
B.subt.	2	122	0	0	0	0	1
B.thur.	0	0	124	0	0	1	0
E.coli	1	1	0	119	0	0	4
M.luteus	0	0	0	0	125	0	0
O.ureth.	0	0	1	3	1	120	0
P.poly.	0	2	0	2	0	1	120

Accuracy: 97.7 %

Sensitivity > 95.2 %, Specificity > 99.3 %

Prediction	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
Reference	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
B.brevis	125	0	0	0	0	0	0
B.subt.	1	124	0	0	0	0	0
B.thur.	0	0	125	0	0	0	0
E.coli	0	0	0	124	0	1	0
M.luteus	0	0	0	0	125	0	0
O.ureth.	0	0	0	0	0	125	0
P.poly.	0	0	0	2	0	0	123

Accuracy: 99.5 %

Sensitivity > 98.4 %, Specificity > 99.7 %

Prediction	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
Reference	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
B.brevis	125	0	0	0	0	0	0
B.subt.	1	124	0	0	0	0	0
B.thur.	0	0	125	0	0	0	0
E.coli	0	0	0	123	0	1	1
M.luteus	0	0	0	0	125	0	0
O.ureth.	0	0	0	2	0	123	0
P.poly.	0	2	0	1	0	0	122

Accuracy: 99.1 %

Sensitivity > 97.6 %, Specificity > 99.6 %

Two alternating laser pulses with wavelengths of 280 and 355 nm.
Distance to sample (outdoor): 22 m. iCCD-channels: 1024. [1]

Averaged spectra for each bacteria with boundaries of 2nd and 3rd quartile

Setup #1 test results

Prediction	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
Reference	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
B.brevis	114	3	0	0	0	2	6
B.subt.	3	102	0	5	0	1	14
B.thur.	0	0	123	1	0	1	0
E.coli	2	2	0	114	0	6	1
M.luteus	0	0	0	0	123	1	1
O.ureth.	2	0	0	3	1	119	0
P.poly.	5	6	0	2	0	1	111

Accuracy: 92.1 %

Sensitivity > 81.6 %, Specificity > 97.1 %

Prediction	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
Reference	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
B.brevis	120	1	0	0	0	1	3
B.subt.	0	119	0	0	0	0	6
B.thur.	0	0	124	0	0	1	0
E.coli	0	0	0	121	0	4	0
M.luteus	0	0	0	0	123	0	2
O.ureth.	2	0	0	0	0	123	0
P.poly.	4	2	0	5	0	0	114

Accuracy: 96.5 %

Sensitivity > 91.2 %, Specificity > 98.5 %

Prediction	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
Reference	B.brevis	B.subt.	B.thur.	E.coli	M.luteus	O.ureth.	P.poly.
B.brevis	114	4	0	0	0	5	2
B.subt.	2	112	0	2	2	2	5
B.thur.	0	0	125	0	0	0	0
E.coli	3	0	0	113	0	5	4
M.luteus	1	2	0	0	121	0	1
O.ureth.	1	0	0	3	0	120	1
P.poly.	6	8	0	4	0	2	105

Accuracy: 92.6 %

Sensitivity > 84.0 %, Specificity > 98.1 %

Prediction	Positive	Negative
Reference	Positive	Negative
Positive	True Positives (TP)	False Negatives (FN)
Negative	False Positives (FP)	True Negatives (TN)

Accuracy = Correct Prediction Rate

$$\frac{TP + TN}{TP + FP + TN + FN}$$

Sensitivity = True Positive Rate

$$\frac{TP}{TP + FN}$$

Specificity = True Negative Rate

$$\frac{TN}{TN + FP}$$

Decision Trees

- Features linearly separated
- Reordered by information entropy
- Minimal number of cases in final nodes prevents overfitting
- Pruning lowers dimensionality

Decision Tree to separate the classes

SVM model for two classes
focused on two features

Support Vector Machines

- Transformation: radial basis function
- Support vectors define the borders
- Linear separation in hyperspace
- Back transformation generates nonlinear model

$$e^{-\sigma \|x - x'\|^2}$$

Artificial Neural Networks

- Features (input) linearly combined
- Choose number of hidden units
- Activation function: Sigmoid
- Weighted connections
- Convergence in most cases

$$\frac{1}{1 + e^{-x}}$$

Neural Network with ten features and
weighted connections after 1000 iterations