scientific reports

Check for updates

OPEN The benefits of haptic feedback in robot assisted surgery and their moderators: a meta-analysis

Max Bergholz^{1,2}, Manuel Ferle¹ & Bernhard M. Weber²

Robot assisted surgery (RAS) provides medical practitioners with valuable tools, decreasing strain during surgery and leading to better patient outcomes. While the loss of haptic sensation is a commonly cited disadvantage of RAS, new systems aim to address this problem by providing artificial haptic feedback. N = 56 papers that compared robotic surgery systems with and without haptic feedback were analyzed to quantify the performance benefits of restoring the haptic modality. Additionally, this study identifies factors moderating the effect of restoring haptic sensation. Overall results showed haptic feedback was effective in reducing average forces (Hedges' g = 0.83) and peak forces (Hedges' g = 0.69) applied during surgery, as well as reducing the completion time (Hedges' q = 0.83). Haptic feedback has also been found to lead to higher accuracy (Hedges' q = 1.50) and success rates (Hedges' g = 0.80) during surgical tasks. Effect sizes on several measures varied between tasks, the type of provided feedback, and the subjects' levels of surgical expertise, with higher levels of expertise generally associated with smaller effect sizes. No significant differences were found between virtual fixtures and rendering contact forces. Implications for future research are discussed.

Robot assisted surgery (RAS) has shown the potential to lead to preferable clinical outcomes for patients¹ while reducing the surgeon's cognitive² and physical workload³. Robot assistance has contributed to improving surgical interventions, especially in the field of minimally invasive surgery. Consequently, robotic surgery systems have seen broad adoption. For example, the majority of prostatectomies performed in the United States in 2010 utilized robotic assistance⁴.

An obvious drawback of RAS is the loss of haptic sensation, depriving surgeons of a natural source of information about interaction forces⁵. This shortcoming has been linked to diminished surgical performance⁶. Several current research and development programs try to rectify this drawback by artificially restoring haptic sensation to the surgeon using master-slave robotic systems⁷⁻⁹. The proposed solutions vary greatly in terms of the methods, how haptic feedback is provided to the surgeon, and their intended fields of application¹⁰⁻¹². A general distinction can be made between direct haptic feedback and haptic sensory substitution. The latter provides information on forces via auditory¹³ or visual¹⁴ cues, whereas direct haptic feedback, which will be the subject of this study, aims to recreate haptic impressions as naturally as possible. The most common approach is kinesthetic haptic feedback, wherein the impedance incurred on the slave side of the robotic system is mirrored back to the master side. For example, Talasaz, Trejos, and Patel¹⁵ used a motorized master controller to mirror the impedance forces picked up by sensors in the robot's end effectors during a robot assisted suturing task. This resulted in a reduction of the forces the subjects applied and thereby lowered risks of suture breakage. Alternatively, haptic feedback is provided via vibrations of the master device¹⁶ or via actuators stimulating the user's fingertips¹⁷. This way, the illusion of holding an object in hand is created.

RAS is most studied in the context of laparoscopic surgery¹⁸, but other paradigms also increasingly employ robotic assistance. In retina surgery, for example, robotic assistance is utilized for the high accuracy and tremorfree stability a robotic end-effector provides. The feedback force can be upscaled to allow the surgeon to sense interaction forces below the human perception threshold¹⁹. Another motivation for using robots in the operating room is to shield the surgeon from harm. In venous catheterization procedures, for example, x-rays are commonly used for live imaging of the patient's vascular system. Haptic feedback allows the surgeon to stay outside the operation room while retaining the haptic sensation crucial for preventing vascular puncturing¹²

While the benefits of haptic feedback technology seem obvious, they must be set against the incurring costs, not only monetary but also in terms of the required additional space and maintenance procedures²⁰. Some researchers also suggest that with adequate expertise, surgeons can extract sufficient information about

¹Department of Ergonomics, Technical University of Munich, 85748 Garching, Germany. ²Institute of Robotics and Mechatronics, German Aerospace Center, 82234 Wessling, Germany. Zemail: manuel.ferle@tum.de

interaction forces during surgery from visual feedback alone. The advanced stereoscopic view of modern robotic surgery systems might allow robotic specialists to gauge interaction forces from visible tissue deformation. It can even create a faux haptic sensation, potentially rendering true haptic feedback redundant²¹. Conversely, other researchers found benefits of haptic feedback even for very experienced surgeons⁸.

It is, therefore, pivotal to provide clinics with information about the expected benefits of haptic feedback systems to allow for informed decisions and to provide researchers with information on which approaches to haptic feedback are the most promising. Existing reviews are primarily qualitative in nature^{22,23}, focused on a single system²⁴ or no longer reflect the current state of development²⁵. Given the field's broad and rapidly evolving nature, an expansive systematic analysis of recent research is needed. In this meta-analysis, key performance metrics were identified, and the overall effects of haptic feedback on these metrics and potential moderating factors were determined. Hence, the present study aims not just to quantify the general benefits of haptic feedback but also their extent under specific conditions, namely, given a particular surgical task, level of users' expertise, and type of feedback. By identifying conditions under which haptic interaction technology is particularly useful, this study provides a basis for decision-making for practitioners and researchers.

Methods

Report compilation

This meta-analysis was performed following the guidelines laid out in the PRISMA statement²⁶. Literature was compiled using several online data libraries. The databases PubMed, IEEEXplore, and Scopus were consulted. This selection has been chosen to cover both key disciplines relevant to the subject, namely medicine (PubMed) and engineering (IEEEXplore), as well as a more generalized database in the form of Scopus to additionally cover applicable papers outside these fields. Given the enormous diversity of paradigms and measures in the field, the search aimed to generate a large number of results rather than trying to minimize the number of irrelevant results. Hence, a broad search string was chosen to compile an as complete as possible body of studies.

(telerobotics OR robot assisted OR telesurgery) AND (haptics OR force feedback OR tactile)

The reports were compiled in September and October 2022. Papers published since 2013 were eligible, the last point in time covered in the most recent meta-analysis on haptic feedback in general RAS²⁵. Additional criteria were defined to exclude non-peer-reviewed publications and to reduce the amount of overlap. This process yielded 1637 papers on PubMed, 1783 on IEEEXplore, and 1630 on Scopus, respectively. 996 of these 5050 results were removed as duplicates. The remaining 4054 papers were screened and filtered by the primary author (Fig. 1).

The following inclusion criteria were defined. The study needs to (1) be peer-reviewed, (2) report quantitative data, (3) feature a direct comparison of subjects' performance with and without haptic feedback, (4) use a master–slave robot system with direct control, (5) be applicable in surgical contexts, (6) make use of direct haptic feedback, (7) be written in English or German. Notably, this excluded papers investigating sensory substitution²⁷, haptic feedback in non-surgical contexts²⁸, or the benefits of haptic feedback for training²⁹. Filtering by these criteria eliminated most search results, leaving N = 115 papers, all of which could be retrieved in full and were consequently subjected to further scrutiny.

The following quality criteria were defined: (1) At least n = 2 subjects per study group, (2) The paper had to account for potential learning effects by counterbalancing or randomizing the order of trials, and (3) No authors participated as test subjects. Nine papers had to be excluded as they failed to meet these criteria. Additionally, 14 papers had been excluded because, upon closer inspection, their compliance with inclusion criteria became dubious. For example, technologically immature systems that would require substantial modification before surgical application were excluded³⁰. 36 papers had to be excluded because of lacking data reporting. The final sample consisted of 56 primary studies with n = 768 individual subjects producing k = 174 observations. A complete list of included primary studies can be found in Table 1.

Data extraction

To calculate effect sizes, means and standard deviations had to be extracted. Where available, these metrics were taken directly from the publications. Otherwise, they were retrieved by the primary author by measuring graphics and figures provided in the primary studies using open-source image manipulation software (GIMP 2.10). This method ensured accuracy on the level of a single pixel and, thereby, the numerical value presented by a pixel. Where this was not possible either, effect sizes were estimated based on sample size and p-values³¹. When papers failed to report precise p-values, the authors of the primary studies were contacted with a request for the data measures in question. If the necessary data could not be retrieved this way either, the study had to be excluded.

Statistical methods

All statistical analyses were performed in "Comprehensive Meta-Analysis Version 4" (CMA, Biostat Inc, Englewood, New Jersey, USA). The following performance metrics have been extracted from the primary studies: (1) Force applied (further divided into average force and peak force), with smaller forces being considered favorable as high forces are associated with increased tissue damage³². (2) Time required, with shorter times being considered desirable. (3) Accuracy, to gauge precision, several sub-measures were combined: deviation from a target angle, deviation from a target point, and deviation from a target path. Lower deviation is desirable; (4) Success rates; in the present data set, this measure chiefly refers to correct tissue identification in palpation tasks.

Since it can be expected that true effect sizes vary between primary studies, a random effects model was chosen. All measurements have been valence-coded such that higher values indicate more desirable outcomes (i.e., shorter times, lower applied forces, less deviation, and higher success rates). The cutoff for statistical significance



Figure 1. Flowchart visualizing the process of sample compilation.

was set to p = 0.05. Hedges' g was chosen to calculate effect sizes rather than the more common Cohen's d since the latter tends to overestimate the effect size, especially for small sample sizes³¹. Hedges' g can thus be seen as more conservative as it adjusts Cohen's d with the factor J.

$$J = 1 - \frac{3}{4df - 1}$$

The size of effects can be interpreted analogously to Cohen's d. That is, effect sizes larger than 0.8 are considered large. They are immediately evident to observers and practitioners and make a meaningful difference in clinical practice. On the other hand, effect sizes below 0.2 are considered irrelevant even if they are statistically significant³¹.

Heterogeneity was assessed with the Q-statistic, I^2 -statistic, and the prediction interval. Following recommendations by Medina et al.³³ it is essential to differentiate these measures. The Q-statistic functions analogously to the F-statistic in primary studies in that it tests against the null hypothesis that all effects in the sample are of equal size. A significant Q-value indicates that not all observed variety between effect sizes results from random error, meaning the effect sizes are heterogeneous. Furthermore, the Q-statistic can be used to test whether the effect sizes observed in different subgroups significantly differ. I^2 quantifies which percentage of the observed variety represents the heterogeneity of the true effect sizes. Lastly, the prediction interval (PI) gives the range of effect sizes 95% of individual effects in a comparable population are expected to fall into. It is similar to, but has to be differentiated from, the confidence interval (CI), which estimates the range the median effect falls into. The confidence interval is reported as well. In this study, the conventional 95% confidence interval was chosen.

First author	Year	Obser-vations	Measures	Subjects' expertise	Task	Surgical field	Feedback type	VF/CF
Abiri ³⁴	2017	1	Time	Inexp	Grasping	Lapa	Cut	CF
Abiri ³⁵	2019a	4	Success, Time	Inexp	Palpation	Lapa	Kin, VT	CF
Abiri ³⁶	2019b	10	Avg F, Peak F	Exp, nov	Grasping	Lapa	Com, Cut, Kin	CF
Abiri ⁷	2019c	1	Point Δ	Nov	Suture	Lapa	VT	CF
Aggravi ³⁷	2021	6	Success, point Δ , time	Inexp	Insertion	Lapa	Kin, VT	CF
Allebas ³⁸	2017	1	Avg F	Mix	Palpation	Lapa	Kin	CF
Bahar ³⁹	2020	6	Deg ∆, Peak F, time	n/a	Insertion	Lapa	Kin	CF, VF
Bao ⁴⁰	2018	1	Time	Exp	Catheter	Cardio	Kin	CF
Camara ⁴¹	2018	1	Success	Inexp	Palpation	Onco	Kin	CF
Chauhan ⁴²	2017	1	Avg F	n/a	Grasping	Lapa	Kin	CF
Chinello ⁴³	2020	6	Avg F, point Δ , time	Inexp	Palpation	Lapa	Com, cut	CF
Chowriappa ⁴⁴	2013	2	Peak F, point Δ	n/a	Insertion	Lapa	Kin	CF
Currie ⁴⁵	2016	1	peak F	Mix	Suturing	Lapa	Kin	CF
Dagnino ⁴⁶	2018	3	Avg F, peak F, time	Inexp	Catheter	Cardio	Kin	CF
Dai ⁴⁷	2019	3	Avg F, point Δ . time	Nov	Suturing	Lapa	Kin	CF
Dalvand ⁴⁸	2014	1	Success	Inexp	Palpation	Lapa	Kin	CF
Diez ⁸	2019	3	Avg F, peak F, time	Exp	Grasping	Lapa	Kin	CF
Ebrahimi ⁴⁹	2018	2	Avg F, time	Mixed	Tracing	Ophtha	Kin	CF
Ehrampoosh ¹⁰	2013	3	Success	n/a	Palpation	Lapa	Kin	CF
Elayaperumal ¹⁰	2014	1	Success	Inexp	Insertion	Lapa	Kin	CF
Fichera ⁵⁰	2015	2	Point Δ	Inexp	Laser	Tors	Com, Kin	VF
Francone ⁵¹	2019	2	Avg F, time	Mixed	Grasping	Lapa	Kin	CF
Gambaro ⁵²	2014	2	Path Δ , time	n/a	Tracing	Lapa	VT	VF
Gerena ⁵³	2020	1	Success	n/a	Grasping	Lapa	Kin	VF
Gibo ⁵⁴	2014	1	Avg F	Inexp	Grasping	Lapa	Kin	CF
Gramma- tikopoulou ⁵⁵	2016	3	Path Δ , success	Novice	Grasping, scan	Lapa	Kin	VF
Howard ¹⁶	2016	1	Point Δ	Novice	Suturing	Lapa	VT	CF
Jeong ⁵⁶	2021	6	Avg F, peak F, time	Inexp	Tracing	Ophtha	Kin	CF
Jin ⁵⁷	2021	2	Avg F, time	Inexp	Catheter	Cardio	Kin	CF
Jou ¹⁷	2022	12	Avg F, Peak F	Exp, mix, nov	Section, grasping	Lapa	Cut	CF
Karponis ⁵⁸	2019	1	Point Δ	Nov	Insertion	Lapa	Kin	CF
Kim ¹¹	2016	1	Peak F	n/a	Grasping	Lapa	Kin	CF
Li ⁹	2022	2	Avg F	Mixed	Catheter	Cardio	Kin	CF
Lim ⁵⁹	2015	3	Avg F	Inexp	Grasping	Lapa	Kin, cut	CF
Lopez ⁶⁰	2013	2	Path Δ , point Δ	n/a	Insertion	Lapa	Kin	VF
Mendelsohn ⁶¹	2020	1	Time	Nov	Section	TORS	VT	CF
Molinero ⁶²	2019	1	Time	Exp	Catheter	Cardio	Kin	VF
Nakazawa ⁶³	2016	2	Point Δ , time	Inexp	Tracing	Neuro	Kin	VF
Navkar ⁶⁴	2013	1	Time	Inexp	Catheter	Cardio	Kin	VF
Olivieri ⁶⁵	2018	2	Path Δ , time	Exp	Laser	TORS	Kin	VF
Ouyang ⁶⁶	2021	3	Avg F, success, time	Mixed	Palpation	Lapa	Cut	CF
Pacchierotti ⁶⁷	2015	6	Avg F, time	Inexp	Grasping	Lapa	Com, Cut, Kin	CF
Pacchierotti ⁶⁸	2016	4	Deg Δ , time	Inexp	Palpation	Lapa	Cut, VT	CF
Portolés ⁶⁹	2015	2	Avg F, time	Inexp	Tracing	Lapa	Kin	CF
Power ⁷⁰	2015	2	Time	n/a	Grasping	Lapa	Kin	CF
Quek ⁷¹	2019	6	Avg F, peak F	Inexp	Grasping	Lapa	Cut, Kin	
Saracino ⁷²	2019	24	Avg F, peak F, time	Exp, inexp, nov	Grasping, palpation, section	Lapa	Kin	CF
Saracino ⁷³	2020	3	Peak F, success, time	Nov	Palpation	Lapa	Kin	CF
Seung ⁷⁴	2017	3	Path Δ	Nov	Tracing, section	Neuro	Kin	VF
Song ⁷⁵	2017	2	Avg F, time	Inexp	Catheter	Cardio	Kin	CF
Tahir ⁷⁶	2022	2	Time	Inexp	Catheter	Cardio	Kin, VT	VF
Talasaz ¹⁵	2017	4	Peak F, time	Mix	Insertion, suture	Lapa	Kin	CF
Urias ⁷⁷	2020	1	Avg F	Exp	Section	Ophtha	Kin	CF
Wottawa ³²	2016	4	Avg F	Exp, nov	Grasping	Lapa	Cut	CF
Continued					-		1	

First author	Year	Obser-vations	Measures	Subjects' expertise	Task	Surgical field	Feedback type	VF/CF
Yin ⁷⁸	2018	1	Time	Mix	Catheter	Cardio	Kin	CF
Zhang ¹²	2021	2	Peak F, time	Mix	Catheter	Cardio	Kin	CF

Table 1. Summary list of studies included in the meta-analysis. "peak F", peak forces; "avg F", average forces. " Δ ", deviation; "exp", experienced subj.; "nov", novice subj.; "inexp", inexperiecned subj. "mixed", diffent levels of expertise in the same group; "n/a", level of expertise not reported; "cardio", cardiovascular; "lapa", laparoscopy; "neuro", neurosurgery; "onco", oncology; "ophtha", ophthalmology; ""TORS", transoral robotic surgery. "com", combined feedback; "cut", cutaneous feedback; "kin", kinesthetic feedback; "VT", vibrotactile feedback. "VF", Virtual Fixtures; "CF", Contact Forces. Mixed and non-available levels of expertise and success rate outcomes were not considered for subgroup analysis.

Subgroup analysis

Based on the reviewed literature, the following categorical variables have been determined as potential moderators and extracted: Level of subjects' expertise, experimental task, type of haptic feedback, and the distinction between virtual fixtures and contact forces. These potential moderators will be briefly described in the following:

Level of expertise

(1) "Inexperienced" denotes subjects with no prior surgical training. (2) "Novices" denotes subjects with medical training but no or limited, i.e., no more than a year of practical experience in the tested procedure. (3) "Experts" denotes subjects with extensive multiyear expertise on the specific procedure tested in the experiment.

Task

Since surgeries often consist of various tasks that partly overlap between procedures, the investigated surgeries were broken down into more elemental tasks for this meta-analysis. The following tasks were extracted from the reviewed body of literature: (1) Catheterization: insertion of venous catheters into the vascular system; (2) Grasping: tasks consisting of grasping, holding, or pulling primarily of tissue, (3) Insertion: puncturing of tissue with needles and/or trocars, (4) Laser: cutting and/or sectioning of tissue involving surgical lasers, (5) Palpation: physical Investigation of tissue with the robot's end effector, (6) Scan: Investigation of tissue via additional means such as ultrasound scans, (7) Section: cutting and/or carving of tissue, (8) Suturing: knot-tying and stitching of tissue, and (9) Tracing: following a line on the surface of a tissue sample.

Feedback type

The different types of feedback employed by researchers to restore haptic sensation have been categorized as follows: (1) Kinesthetic feedback: application of a force vector to the master side of the robot mirroring the impedance offered by an object in contact with the slave side. (2) Vibrotactile feedback: provision of information to the user via the vibration of either a master controller or an externally worn device. (3) Cutaneous feedback: provision of haptics via the motion of actuators against the operator's fingertips, creating an illusion of holding an object. (4) Combined: any combination of the above.

Contact vs. virtual fixtures

(1) Contact forces: Feedback is given about contact with physical objects at the robotic site. (2) Virtual fixture: Feedback is given in response to virtual points, which have been defined pre-operatively in relation to either the patient's body (e.g., a virtual point close to sensible tissue) or the posture of the robot (e.g., a certain degree on the slave-robot's joints).

In addition to the moderators, we list the relevant surgical fields in the table below. While several studies are potentially applicable to multiple fields, an intended field of application can be noted for each paper. The following surgical fields were considered in the primary studies: (1) Cardiovascular surgery, (2) Laparoscopic surgery, (3) Neurosurgery, (4) Oncology, (5) Ophthalmology, and (6) Oncology.

Results

Figures 2, 3 and 4 show the effect sizes, their variances, and overall prediction and confidence intervals. Forest plots have been created in CMA. Note that effect sizes in the plots are natively rather than valence-coded due to software limitations.

Effects on applied force

Main effects

<u>Average forces.</u> Based on the analysis of k = 52 study groups comprised of a total of n = 249 individuals, a strong effect of haptic feedback on the average forces was found (Hedges' g = 0.83, CI 0.63–1.03). The Q-test indicates that effect sizes varied substantially between studies (Q(51) = 135.95, p < 0.001). 62% of that difference in effect size is accounted for by the heterogeneity of the true effect rather than random error (I^2 = 0.62).

<u>Peak force.</u> K=37 study groups with a total of n=172 individual subjects were analyzed. A moderate effect on peak forces was found (g=0.69, CI 0.51–0.87). True effect sizes differed in the population (Q(36)=51.7, p=0.043). I^2 = 30% of the observed difference can be attributed to the difference in true effect size.

5

Study name	Subgroup within study Outcome		Statistics	for each	study		Hedges's g and 95% Cl
		Hedges's g	Variance	Lower limit	Upper limit	p-Value	
Song et al. 2017	1 force avg	-3,28	0,62	-4,82	-1,73	0,00	
Li et al. 2022	1 force avg	-2,44	0,53	-3,87	-1,01	0,00	
Allebas et al. 2017	1 force avg	-2,39	0,27	-3,42	-1,37	0,00	
Pacchierotti et al. 2015	1 force avg	-2,31	0,22	-3,22	-1,40	0,00	
Lim et al. 2015	1 force avg	-2,18	0,33	-3,31	-1,05	0,00	
Cauhan et al. 2017	1 force avg	-2,17	0,51	-3,57	-0,77	0,00	
Pacchierotti et al. 2015	3 force avg	-2,16	0,20	-3,05	-1,28	0,00	
Lim et al. 2015	2 force avg	-1,97	0,31	-3,06	-0,88	0,00	
Abiri, Pensa et al. 2019	4 Combined	-1,69	0,23	-2,62	-0,76	0,00	
Francone et al 2019	1 force avg	-1,69	0,40	-2,93	-0,44	0,01	
Zhang et al. 2021	1 force peak	-1,62	0,25	-2,59	-0,64	0,00	
Jeong et al. 2021	1 Combined	-1,55	0,25	-2,53	-0,57	0,00	
Quek et al. 2019	3 Combined	-1,46	0,24	-2,41	-0,50	0,00	
Kim et al. 2016	1 force peak	-1,45	0,19	-2,30	-0,60	0,00	
Quek et al. 2019	1 Combined	-1,44	0,24	-2,39	-0,49	0,00	
Lim et al. 2015	3 force avg	-1,37	0,25	-2,36	-0,38	0,01	
Abiri, Pensa et al. 2019	1 Combined	-1,36	0,16	-2,14	-0,58	0,00	
Chowriappa et al. 2013	1 force peak	-1,26	0,41	-2,51	-0,01	0,05	
Talasatz et al. 2017	1 force peak	-1,16	0,14	-1,89	-0,43	0,00	
Currie et al. 2016	1 force peak	-1,11	0,18	-1,95	-0,28	0,01	
Jeong et al. 2021	2 Combined	-1,10	0,18	-1,93	-0,28	0,01	
Abiri, Pensa et al. 2019	3 Combined	-1,06	0,14	-1,80	-0,31	0,01	
Pacchierotti et al. 2015	2 force avg	-1,05	0,14	-1,79	-0,30	0,01	
Jin et al. 2021	1 force avg	-1,02	0,38	-2,23	0,18	0,10	
Dai et al. 2019	1 force avg	-0,99	0,13	-1,69	-0,29	0,01	
Dagnino et al. 2018	1 Combined	-0,96	0,21	-1,85	-0,06	0,04	
Chinello et al. 2020	2 force avg	-0,83	0,06	-1,33	-0,34	0,00	
Talasatz et al. 2017	2 force peak	-0,83	0,13	-1,53	-0,12	0,02	
Ebrahimi et al. 2018	1 force avg	-0,83	0,46	-2,16	0,51	0,22	
Diez et al. 2019	1 Combined	-0,82	0,22	-1,74	0,10	0,08	
Urias et al. 2020	1 force avg	-0,77	0,40	-2,01	0,47	0,22	
Saracino et al. 2020	1 force peak	-0,74	0,09	-1,34	-0,14	0,02	
Bahar et al. 2020	1 force peak	-0,72	0,20	-1,59	0,15	0,10	
Saracino et al. 2019	8 Combined	-0,69	0,35	-1,85	0,47	0,24	
Quek et al. 2019	4 Combined	-0,65	0,19	-1,51	0,22	0,14	
Chinello et al. 2020	1 force avg	-0,57	0,05	-1,02	-0,11	0,01	
Ouyang et al. 2021	I force avg	-0,54	0,04	-0,95	-0,13	0,01	
Abiri, Pensa et al. 2019	3 Combined	-0,55	0,13	-1,24	0,17	0,14	
Abiri, Pensa et al. 2019	2 Combined	-0,52	0,13	-1,22	0,17	0,14	
Nottawa et al 2016	4 Torce avg	-0,46	0,07	-0,99	0,06	0,08	
Sanar et al. 2020	2 Iorce peak	0,44	0,19	-1,29	0,41	0,51	
Gibo et al. 2019	5 Combined	-0,39	0,33	-1,53	0,74	0.50	
Saracino et al. 2019	11 Combined	-0,38	0.35	-1,22	0,47	0,50	
Saracino et al. 2019	4 Combined	-0,37	0,35	-1,50	0,78	0.52	
Saracino et al 2019	5 Combined	-0,32	0,55	-1 40	0,02	0.64	
Wottawa et al 2016	3 force and	-0,27	0,55	_0.7⊑	0,05	0.33	
lou et al. 2022	5 Combined	-0,25	0,07	-1 30	0,23	0.67	
Jou et al. 2022	2 Combined	-0.24	0.44	-1.53	1.06	0.72	
Jou et al. 2022	4 Combined	-0.18	0.33	-1.30	0.95	0.76	
Jou et al. 2022	6 Combined	-0.07	0.33	-1.20	1.05	0.90	
Saracino et al. 2019	1 Combined	-0.07	0.33	-1.20	1,06	0.90	
Jou et al. 2022	3 Combined	-0.07	0.33	-1.19	1.05	0.91	
Saracino et al. 2019	2 Combined	-0.05	0.33	-1.17	1.07	0.93	
Saracino et al. 2019	6 Combined	-0,04	0,33	-1,16	1,08	0,95	
Portolles et al. 2015	1 force ave	0,10	0,18	-0,74	0,94	0,82	
Jou et al. 2022	1 Combined	0,13	0,43	-1,15	1,41	0,84	
Wottawa et al 2016	2 force ave	0.14	0.13	-0.56	0.85	0.69	
Quek et al. 2019	2 Combined	0.14	0.18	-0.70	0.99	0.74	
Saracino et al. 2019	7 Combined	0.46	0.34	-0.68	1.60	0.43	
Wottawa et al 2016	1 force ava	0.83	0.20	-0,04	1,70	0,06	
Pooled		-0,83	0,01	-1,01	-0,66	0,00	
Prediction Interval		-0.83		-1.86	0.19		
		0,00		1,00	5,25		

Figure 2. Effects on haptic feedback on applied force.

<u>Combined force.</u> The effect size for combined average and maximum applied forces was also calculated. For study groups that reported both outcomes, combined means were used in line with the recommendations of

Lower forces

Higher Forces

itudy name	Subgroup within study Outcome			Statistics	for each	study		Hedges's g and 95	
			Hedges's g	Variance	Lower limit	Upper limit	p-Value		
ong et al. 2017	1	time	-4,34	1,03	-6,33	-2,36	0,00	⊨● ∔	
Bao et al. 2018	1	time	-3,77	0,54	-5,21	-2,33	0,00		
Abiri et al. 2019	2	time	-3,60	0,27	-4,62	-2,58	0,00		
'in et al. 2018	1	time	-3,28	0,62	-4,82	-1,73	0,00		
acchierotti et al. 2015	1	time	-2,89	0,27	-3,90	-1,88	0,00		
i et al. 2022	1	time	-2,68	0,58	-4,17	-1,18	0,00		
ahir et al. 2022	2	time	-2,40	0,33	-3,52	-1,28	0,00		
ahir et al. 2022	1	time	-2,25	0,31	-3,34	-1,16	0,00		
acchierotti et al. 2015	3	time	-2,21	0,21	-3,10	-1,31	0,00		
in et al. 2021	1	time	-2,05	0,54	-3,49	-0,62	0,01		
hang et al. 2021	1	time	-1,98	0,28	-3,01	-0,94	0,00		
Abiri & al. 2016	1	time	-1,81	0,18	-2,64	-0,97	0,00		
rancone et al 2019	1	time	-1.47	0.37	-2.67	-0.27	0.02		
Duvang et al. 2021	1	time	-1.14	0.06	-1.63	-0.64	0.00		
Aggravi et al. 2021	2	time	-1.10	0.17	-1.91	-0.30	0.01		
Pacchierotti et al. 2016	1	time	-1.07	0.12	-1.75	-0.38	0.00		
Javkar et al. 2013	1	time	-0.99	0.32	-2 19	0 21	0 11		
Pacchierotti et al. 2015	2	time	-0.98	0.14	-1 72	-0.25	0.01		
acchierotti et al. 2015	2	time	-0,96	0.12	-1.63	-0.28	0,01		
Portellos et al. 2015	1	time	0,50	0,12	1.04	0,20	0,01		
biri et al. 2010	1	time	-0,95	0,21	1 46	-0,00	0,04		
course at al. 2019	1	time	-0,82	0,11	1 77	-0,17	0,01		
	1	time	-0,80	0,24	-1,//	0,17	0,11		
eong et al. 2021	1	time	-0,72	0,13	-1,43	-0,01	0,05		
alasatz et al. 2017	1	time	-0,63	0,13	-1,32	0,06	0,07		
Aggravi et al. 2021	1	time	-0,63	0,15	-1,39	0,13	0,11		
alasatz et al. 2017	2	time	-0,62	0,12	-1,31	0,07	0,08		
Dagnino et al. 2018	1	time	-0,59	0,19	-1,45	0,27	0,18		
chinello et al. 2020	2	time	-0,59	0,05	-1,05	-0,13	0,01		
chinello et al. 2020	1	time	-0,48	0,05	-0,93	-0,04	0,03		
ower et al. 2015	2	time	-0,43	0,23	-1,37	0,51	0,37		
lakazawa et al. 2016	1	time	-0,32	0,33	-1,45	0,81	0,58		
eong et al. 2021	2	time	-0,20	0,10	-0,82	0,43	0,54		
livieri et al. 2018	2	time	-0,19	0,11	-0,84	0,47	0,58		
ahar et al. 2020	2	time	-0,17	0,18	-1,01	0,67	0,69		
livieri et al. 2018	1	time	-0,11	0,11	-0,76	0,55	0,75		
aracino et al. 2019	7	time	-0,10	0,33	-1,22	1,02	0,86		
ahar et al. 2020	1	time	0,01	0,18	-0,83	0,84	0,99		
aracino et al. 2020	1	time	0,04	0,09	-0,54	0,62	0,89	♦	
aracino et al. 2019	4	time	0,07	0,33	-1,05	1,19	0,91		
iez et al. 2019	1	time	0,13	0,20	-0,75	1,01	0,78		
aracino et al. 2019	2	time	0,14	0,33	-0,98	1,26	0,80		
ambaro et al. 2014	1	time	0,20	0,16	-0,57	0,98	0,61		
1endelssohn et al. 2020	1	time	0,20	0,18	-0,64	1,04	0,64		
ai et al. 2019	1	time	0,23	0,05	-0,23	0,69	0,32		
aracino et al. 2019	6	time	0,28	0,33	-0,84	1,41	0,62		
1olinero et al. 2019	1	time	0,42	0,44	-0,88	1,72	0,53		
aracino et al. 2019	11	time	0,53	0,34	-0,61	1,67	0,36		
aracino et al. 2019	9	time	0,94	0,37	-0,25	2,14	0,12		
brahimi et al. 2018	1	time	1,54	0,58	0,04	3,03	0,04		
ooled			-0,83	0,02	-1,09	-0,57	0,00		
			-,				,		

Lower completion time

Figure 3. Effects on haptic feedback on completion time.

Borenstein et al.³¹. k=61 Study groups were analyzed. The overall effect on force was large (g=0.83, CI 0.66–1.01). True effect sizes in the population varied (Q(60)=147.20, p<0.001), and this variance accounts for I^2 =59% of the observed variance.

Moderation

Only combined force was considered for moderation effects to achieve a reasonable number of observations per cell. Only 46 study groups could be analyzed for expertise level due to lacking data reports in the primary studies.

7

6,00

Higher completion tim

udy name	Subgroup within study	Outcome		Statistics	for each	study	
			Hedges's g	Variance	Lower limit	Upper limit	p-Value
Aggravi et al. 2021	2	point deviation	-5,81	0,79	-7,55	-4,06	0,00
Olivieri et al. 2018	1	path deviation	-3,90	0,34	-5,04	-2,77	0,00
Aggravi et al. 2021	1	point deviation	-3,88	0,43	-5,17	-2,59	0,00
Abiri et al. 2020	1	point deviation	-3,16	0,29	-4,22	-2,10	0,00
Fichera et al. 2015	1	point deviation	-2,27	0,31	-3,36	-1,17	0,00
Pacchierotti et al. 2016	1	deg deviation	-2,22	0,17	-3,03	-1,40	0,00
Seung et al. 2017	3	path deviation	-2,13	0,66	-3,73	-0,54	0,01
Chowriappa et al. 2013	1	point deviation	-2,10	0,55	-3,55	-0,65	0,00
Pacchierotti et al. 2016	2	deg deviation	-1,92	0,16	-2,70	-1,14	0,00
Karponis et al. 2019	1	point deviation	-1,85	0,76	-3,56	-0,15	0,03
Olivieri et al. 2018	2	path deviation	-1,73	0,16	-2,51	-0,96	0,00
Lopez et al. 2013	1	Combined	-1,60	0,45	-2,91	-0,29	0,02
Seung et al. 2017	2	path deviation	-1,25	0,48	-2,60	0,11	0,07
Grammatikopoulo et al. 2016	3	path deviation	-1,11	0,13	-1,82	-0,39	0,00
Howard & Swewczyk 2016	1	point deviation	-0,88	0,13	-1,58	-0,17	0,02
Fichera et al. 2015	2	point deviation	-0,84	0,20	-1,72	0,04	0,06
Gambaro et al. 2014	1	path deviation	-0,78	0,17	-1,59	0,02	0,06
Grammatikopoulo et al. 2016	2	path deviation	-0,73	0,10	-1,35	-0,10	0,02
Chinello et al. 2020	2	point deviation	-0,70	0,06	-1,18	-0,23	0,00
Chinello et al. 2020	1	point deviation	-0,65	0,06	-1,12	-0,18	0,01
Seung et al. 2017	1	path deviation	-0,52	0,45	-1,83	0,79	0,44
Bahar et al. 2020	2	deg deviation	-0,36	0,19	-1,21	0,49	0,40
Dai et al. 2019	1	point deviation	-0,21	0,05	-0,67	0,25	0,37
Nakazawa et al. 2016	1	point deviation	-0,09	0,33	-1,21	1,03	0,88
Bahar et al. 2020	1	deg deviation	0,17	0,18	-0,67	1,01	0,69
Pooled			-1,50	0,05	-1,92	-1,07	0,00
Prediction Interval			-1,50		-3,51	0,51	

Lower deviation

Higher deviation

Figure 4. Effects of haptic feedback on accuracy.

Significant differences (Q(2) = 13.05, p < 0.001) between levels of subjects' expertise were found. Significant effects were found for inexperienced (g = 0.98) and novice subjects (g = 0.84) but not for experts.

Force data was reported for seven different tasks, and significant effect size differences (Q(6) = 16.9, p < 0.001) emerged, with effect sizes being the largest for catheterization tasks (g = 1.72). Only for section tasks, no significant effect of haptic feedback was found.

No study using vibrotactile feedback reported data for applied forces. Effect sizes were different between subgroups (Q(2) = 17.92, p < 0.001), being largest for combined (g = 1.39) and smallest for cutaneous feedback (g=0.36). As Virtual Fixtures aim to provide haptic feedback before contact occurs, no comparison between Virtual Fixtures and Contact Forces can be made. For a full list of effects, refer to Table 2.

Effects on completion time

Main effects

An analysis of n = 372 subjects across k = 49 study groups found a large effect (g = 0.83, CI 0.57-1.09). As for forces, the effects on time showed a significant heterogeneity (Q(48) = 229.10, p < 0.001), accounting for I^2 = 79% of the observed variance.

Group	Observations	Hedges' g	95% CI	p-value	Q-value
Overall	61	0.83	0.66 to 1.01	< 0.01	
Expertise					
Experienced	10	0.15	-0.18 to 0.47	0.37	
Novice	13	0.68	0.44 to 0.93	< 0.01	
Inexperienced	23	0.98	0.65 to 1.31	< 0.01	
Total between	3			< 0.01	13.05
Task					
Catheter	5	1.72	0.95 to 2.50	< 0.01	
Grasp	29	0.88	0.59 to 1.17	< 0.01]
Insertion	4	0.87	0.43 to 1.31	< 0.01	
Palpation	7	0.70	0.35 to 1.04	< 0.01	
Section	9	0.16	- 0.23 to 0.55	0.42	
Suturing	3	0.96	0.54 to 1.39	< 0.01	
Tracing	4	0.83	0.08 to 1.57	0.03	
Total between	7			0.01	16.92
FB Туре					
Cutaneous	18	0.36	0.11 to 0.61	0.01	
Kinesthetic	37	0.96	0.75 to 1.18	< 0.01	
Combined	6	1.39	0.83 to 1.96	< 0.01	1
Total between	3			< 0.01	17.92

Table 2. Effect size of haptic feedback on applied forces overall and across subgroups. Significant values are in [bold]. Columns from left to right denote the subgroup, the number of observations per subgroup, the effect of providing haptic feedback, the confidence interval of the effect size, and its p-value. The total between the p-value and Q-value indicates whether the effect sizes differ significantly between subgroups.

Moderation

36 observations reported subjects' expertise. Significant effect size differences (Q(2) = 37.46, p < 0.001) between levels of expertise were found, with time reduction only manifesting for inexperienced subjects (g=1.21).

For the eight different tasks, effect sizes differed significantly (Q(7) = 10.98, p < 0.001). Significant effects were found for catheterization (g = 2.07), grasping (g = 1.30), insertion (g = 0.53) and palpation (g = 0.69).

No significant effect size difference between feedback types was found. Similarly, contact force rendering and virtual fixtures provided significant time savings but did not differ significantly. Full results are reported in Table 3.

Effects on accuracy

Main effects

K = 25 study groups with n = 192 individuals were analyzed. Of these 25 observations, 13 had natively reported point-deviation, 7 path-deviation, 4 degree-deviation, and 1 had already natively aggregated these measures (see Fig. 4). An overall effect of g = 1.50 (CI 1.07–1.92) was found. The true effect sizes varied in the population (Q(24) = 145.94, p < 0.001). The variance of the true effect accounted for I^2 = 0.83 of the observed variance.

For 20 study groups, expertise was reported, but no effect size difference between levels of expertise was found. Interestingly, the observed effect was highest for experts (g=2.78), but the sample was very small, with only k=2 expert groups that reported accuracy measures.

Accuracy measures for seven tasks were reported. A significant effect for haptic feedback was found for every task except for suturing. However, the size of the effects did not differ significantly from each other.

Comparing feedback types, significant effects were found for every type except combined feedback. Again, the number of groups in this category was rather small (k = 2). No significant effect size difference between groups was found.

A comparison of the effect sizes between Virtual Fixtures and Contact Forces showed significant effects of haptic feedback for both groups. There was no significant difference in effect sizes. A full list of results can be found in Table 4.

Success rates

Lastly, success rates showed, based on the analysis of k = 15 study groups with n = 181 subjects, a large effect of g = 0.80 (CI 0.24–1.35). Substantial heterogeneity was present (Q(14) = 71.33, p < 0.001), which accounted for $I^2 = 8037\%$ of the observed variance. The prediction i3dnterval is PI – 1.36 to 2.96. Due to the small number of observations for this measure, most of which fall under the same paradigm (i.e., palpation), this measure had to be excluded from subgroup analysis.

Group	Observations	Hedges' g	95% CI	p-value	Q-value
Overall	49	0.83	0.57 to 1.09	< 0.01	
Expertise			·		
Experienced	7	0.38	- 0.37 to 1.12	0.32	
Novice	5	-0.24	- 0.55 to 0.06	0.12	
Inexperienced	24	1.21	0.86 to 1.56	< 0.01	
Total between	3			< 0.01	37.46
Task					
Catheter	11	2.07	1.30 to 2.84	< 0.01	
Grasp	8	1.30	0.62 to 1.97	< 0.01	
Insertion	5	0.53	0.16 to 0.90	< 0.01	
Laser	2	0.15	- 0.32 to 0.61	0.53	
Palpation	11	0.69	0.21 to 1.17	< 0.01	
Section	4	-0.18	- 0.69 to 0.34	0.50	
Suturing	3	0.15	- 0.68 to 0.98	0.72	
Tracing	4	0.22	- 0.31 to 0.74	0.42	
Total between	8			< 0.01	10.98
FB type					
Cutaneous	6	0.97	0.63 to 1.30	< 0.01	
Kinesthetic	35	0.67	0.35 to 0.98	< 0.01	
Vibrotactile	6	1.26	0.18 to 2.34	0.02	
Combined	2	1.69	- 0.56 to 3.94	0.14	
Total between	4			0.42	2.81
VF vs. CF					
Contact Force	38	0.90	0.59 to 1.20	< 0.01	
Virtual Fixture	11	0.59	0.11 to 1.07	< 0.01	
Total between	2			0.30	1.09

 Table 3. Effect size of haptic feedback on completion time across subgroups. Significant values are in [bold].

.....

Publication bias

To assess a potential publication bias, the classic Fail-Safe-N was calculated for force, time, and accuracy measures. It indicates how many studies with an effect size of zero would have to be found for the meta-analysis to no longer be significant on a level of $p < 0.05^{80}$. The test returned a value of N = 9627. This means that 9627 hypothetical unpublished or otherwise missed papers with no effect for haptic feedback had to exist to invalidate the findings of this analysis via publication bias. Although it is important to acknowledge the limitations of this measure, given the large number, potential bias can be deemed too weak to fundamentally alter results. A visual inspection of the funnel plot (see Appendix 1) indicates no strong bias either. If bias is present in the sample, the analysis can be expected to react robustly.

Discussion

To quantify the effect of haptic feedback on surgical performance, the results of 56 primary studies from 2013 onwards were aggregated using meta-analytical methods. Significant overall effects were found for all investigated measures: applied forces, completion time, accuracy, and success rates. However, the present study goes beyond a simple analysis of the main effect. To identify the specific conditions under which the benefits of haptic feedback are largest, the moderating effects of expertise, task, and feedback type were investigated via subgroup analyses. The subject's experience and the specific nature of the task were identified to have the greatest influence on how beneficial haptic feedback is.

Large overall effects were found across paradigms and measures. The effects found in the present study were larger than those described by Weber and Eichberger²⁵ in an earlier meta-analysis. In particular, the present study's effects on completion time (g=0.83 vs. g=0.22) and accuracy (g=1.5 vs. g=0.69) were substantially larger. This is likely to reflect both technological progress and an application of haptic feedback to new surgical procedures (e.g., catheterization tasks). This underscores haptic feedback's continued and expanding role in medical practice.

A strong effect on applied forces was found overall and in most subgroups, with a notable exception being experts. This finding is of critical importance as some research suggests higher rates of patient tissue damage during RAS⁶, which has been linked to higher interaction forces⁷⁹. This may be the most considerable benefit of haptic feedback for robot-assisted procedures. Furthermore, the effects were largest when combined feedback was employed, suggesting it is a fruitful avenue of development.

Effects on completion times were mixed, with no significant effects of haptic feedback for tracing, suturing, and laser surgery tasks. This is also in line with inconsistent findings of previous meta-analyses: Nitsch and

Group	Observations	Hedge's g	95% CI	p-value	Q-value
Overall	25	1.50	1.07 to 1.92	< 0.01	
Expertise					
Experienced	2	2.78	0.65 to 4.90	0.01	
Novice	9	1.20	0.61 to 1.78	< 0.01]
Inexperienced	9	1.87	1.07 to 2.67	< 0.01	
Total between	3			0.20	3.23
Task					
Insertion	7	2.12	0.68 to 3.57	< 0.01	
Laser	4	2.14	0.97 to 3.31	< 0.01	
Palpation	4	1.30	0.57 to 2.03	< 0.01	1
Scan	1	1.11	0.39 to 1.82	< 0.01	
Section	2	1.62	0.59 to 2.65	< 0.01	
Suturing	3	1.34	- 0.11 to 2.80	0.07	
Tracing	4	0.63	0.20 to 1.05	< 0.01	
Total between	7			0.09	10.98
FB type	Observations	Hedges' g	95% CI	p-value	Q-value
Cutaneous	2	1.25	0.00 to 2.49	0.05	
Kinesthetic	16	1.30	0.75 to 1.84	< 0.01	
Vibrotactile	5	2.40	1.08 to 3.73	< 0.01	
Combined	2	1.40	- 0.12 to 2.93	0.07	
Total between	4			0.50	2.39
VF vs. CF					
Contact Force	12	1.76	1.06 to 2.45	< 0.01	
Virtual fixture	13	1.28	0.78 to 1.78	< 0.01	1
Total between	2			0.28	1.19

Table 4. Effect size of haptic feedback on accuracy across subgroups. Significant values are in [bold].

Färber⁸⁰, e.g., found a significant effect for completion time (g=0.75), whereas Weber and Schneider⁸¹ did not report such an effect. This might be because the added guidance gives users the security necessary for some tasks to complete the task quickly. However, for other tasks, this might be canceled out by the haptic feedback, prompting a more cautious behavior. This is corroborated by the higher accuracy and reduced force application found in the present study.

Effects on Accuracy were overall largest and crucially the only metric on which significant effects for experts were found. Due to the very small number of studies in this subgroup (k=2), this finding should be interpreted cautiously.

Overall, the level of expertise emerged as an important moderator, with effects generally smaller for experts. This is consistent with expectations among practitioners and researchers suggesting that long-term practitioners of RAS develop techniques to partially compensate for the loss of haptic via their practice⁸². However, whereas, e.g., Hagen et al.²¹ suggest that even beginning surgeons can compensate for missing haptic feedback by visual cues alone, this meta-analysis found substantial effects for novices. Combined with results suggesting that haptic feedback enhances surgeons' training⁶, it can be constituted that especially less experienced surgeons benefit from haptic feedback systems.

This study found that haptic feedback benefits depend on the task demands, with the strongest effects for catheterization tasks and the weakest effects for sectioning tasks. Catheterization is a relatively new field of applications for haptic feedback. During vascular catheterization, the surgeon has to rely on low-fidelity vision provided by MRT or X-ray scans while performing a task that requires extreme caution to prevent injuring the patient. Based on the strong positive effects on completion time and force application found in this review, it can be predicted that haptic feedback will play a crucial role in this intervention. Especially considering the use of x-ray imaging is a strong motivator for reduced patient exposure times and the fact that instrument contact forces well below one Newton can lead to vascular puncturing⁸³.

Benefits on all measures were found for grasping and palpation tasks. This is an important finding considering the high prevalence of laparoscopic interventions among RAS, where these tasks play a crucial role²². For suturing—another common laparoscopic task—only force regulation strongly benefitted from haptic feedback. This is, however, crucial to prevent the common problem of suture breakage. Many researchers have argued that the widespread adoption of haptic feedback systems would greatly benefit this robot assisted laparoscopy^{4,20}. The present study substantiates this argument.

Insertion tasks showed substantially reduced forces and completion time and improved accuracy when haptic feedback was available. In non-robot-assisted surgery, the shear forces incurred by different layers of tissue are the primary means for surgeons to guide the correct insertion of instruments¹⁰. Artificial haptic feedback restores this method for RAS.

Tracing, a task common in retinal surgery, showed strong effects for force and accuracy but not for time. Given that in retina surgery, time is usually a less critical metric due to the absence of time-constraining circumstances like heavy blood loss, these results are nonetheless encouraging.

For the other tasks (sectioning, laser surgery, and scanning), results were comparatively mixed but still overall positive (except for sectioning tasks). For these tasks, advancements in stereoscopic 3D vision and visual feedback may already provide enough guidance for surgeons, as suggested by the findings of Weber and Schneider⁸⁰.

That no form of haptic feedback emerged clearly superior contrasts with the findings of Weber and Eichberger²⁵, who found larger effect sizes for kinesthetic feedback than for vibrotactile feedback. The studies of the present meta-analysis skewed heavily away from the latter modality, with only k = 11 observations for pure vibrotactile feedback. The analysis by Weber and Eichberger²⁵ included a total of k = 55 observations for this feedback type. This discrepancy in data may explain the different findings. Furthermore, this indicates a trend away from pure vibrotactile feedback in favor of kinesthetic and combined feedback. Future studies could integrate the data of several meta-analyses on the topic to further investigate the differences.

No significant effect size difference was found between haptic feedback and virtual fixtures. There is the potential of an interaction here precluding an effect, as a look into the studies revealed that virtual fixtures were more commonly combined with vibrotactile feedback and force rendering more commonly with kinesthetic feedback (see Table 1).

In fact, a deeper three-way interaction between measure, task, and feedback can be conceived. For example, vibrotactile feedback might be more effective in reducing completion time than kinesthetic feedback, specifically for palpation tasks³¹. The present study, however, lacks the power to test for such interactions.

Another potential limitation of this analysis is that several related sub-measures had to be aggregated as each sub-measure by itself had too few observations per factor level to allow for subgroup analysis. Combining similar outcomes to organize and investigate large data sets is a core purpose of meta-analyses, but in doing so, the comparability of data must be considered³¹. Since peak and average forces just constitute two different data points of the same measurement, comparability can be readily assumed. In the case of deviation, it can be noted that the different sub-measures are inherently linked. During a needle insertion task, the surgeon may have to guide the instrument's tip to an exact location to administer a substance precisely and avoid unnecessary tissue damage³⁶. In this example, deviation from this target point and deviation from the ideal angle of insertion are intrinsically linked, and combining them can, therefore, be considered valid.

Lastly, there is the potential that the image-based retrieval mentioned in the method section, made necessary by lacking direct reporting in the primary studies, is less accurate than direct numerical retrieval. A total of 33 observations in the final study were retrieved this way. To test whether this retrieval method has skewed the data, the effect size of all image-retrieved outcomes was compared to the effect size of all numerically retrieved outcomes. If the image-based retrieval leads to comparable reliability as direct numerical retrieval, no difference in effect sizes should exist. Indeed, no significant difference in effect sizes was found (Q(1)=0.11, p=0.74). It was thus concluded that the retrieval method yielded sufficient reliability, and image-retrieved data was included in the analysis.

Conclusion

With the benefits of haptic feedback firmly established and application fields where its impact is especially strongly identified, research can focus on these fields to investigate further factor interactions and compare different forms of haptic feedback. In particular, a comparison between direct haptic feedback and haptic substitution to identify in which scenarios one is preferable over the other appears to be a fruitful next step (Table 5).

In summary, providing haptic feedback was found to have desirable effects across task and feedback paradigms. The effects are diminished for practitioners with high levels of expertise but remain descriptively present. Reduced interaction forces and completion times have great potential to limit tissue damage and blood loss during the operation and thereby vastly improve patient safety. These findings can serve as a basis to inform equipment acquisition and future research directions.

Key takeaways	
Main effects	Large positive effects of providing haptic feedback were found for all measures: applied force (g=0.83), completion time (g=0.83), accuracy (g=1.50), and success rates (g=0.80)
Level of expertise	Expertise moderates the effects found for applied forces and completion time but not for accuracy. Generally, the difference between haptic and no haptic feedback is larger for less experienced users
Task	The type of task likewise moderates the effects found for forces and completion time but not for accuracy. In both cases, the difference between haptic and no haptic feedback was largest for catheterization and smallest for cutting tasks
Feedback type	How the haptic feedback was delivered only affected the effects found for forces. Cutaneous feedback had the weakest (g =0.36) and combined feedback the strongest effect (g =1.39)
Virtual fixtures vs. contact forces	Whether the feedback was given about actual physical contact or virtual fixtures did not affect effect size

 Table 5.
 Summary of most important results.

.....

Data availability

All data generated during this study is published in this article. All primary studies analyzed in this meta-analysis have been reported and are publicly available.

Received: 3 May 2023; Accepted: 3 November 2023 Published online: 06 November 2023

References

- Wang, X., Cao, G., Mao, W., Lao, W. & He, C. Robot-assisted versus laparoscopic surgery for rectal cancer: A systematic review and meta-analysis. J. Cancer Res. Ther. 16, 979–989. https://doi.org/10.4103/jcrt.JCRT_533_18 (2020).
- Klein, M. I. *et al.* Mental workload and stress perceived by novice operators in the laparoscopic and robotic minimally invasive surgical interfaces. *J. Endour.* 26, 1089–1094. https://doi.org/10.1089/end.2011.0641 (2012).
- Fan, X., Forsman, M., Yang, L., Lind, C. M. & Kjellman, M. Surgeons' physical workload in open surgery versus robot-assisted surgery and nonsurgical tasks. Surg. Endosc. 36, 8178–8194. https://doi.org/10.1007/s00464-022-09256-0 (2022).
- 4. Lowrance, W. T. *et al.* Contemporary open and robotic radical prostatectomy practice patterns among urologists in the United States. *J. Urol.* **187**, 2087–2092. https://doi.org/10.1016/j.juro.2012.01.061 (2012).
- Enayati, N., De Momi, E. & Ferrigno, G. Haptics in robot- assisted surgery: Challenges and benefits. *IEEE Rev. Biomed. Eng.* 9, 49–65. https://doi.org/10.1109/RBME.2016.2538080 (2016).
- van der Meijden, O. A. & Schijven, M. P. The value of haptic feedback in conventional and robot-assisted minimal invasive surgery and virtual reality training: A current review. Surg. Endosc. 23, 1180–1190. https://doi.org/10.1007/s00464-008-0298-x (2009).
- Abiri, A. et al. Suture breakage warning system for robotic surgery. IEEE Trans. Biomed. Eng. 66, 1165–1171. https://doi.org/10. 1109/TBME.2018.2869417 (2019).
- Díez, S. P. et al. Evaluation of haptic feedback on bimanually teleoperated laparoscopy for endometriosis surgery. *IEEE Trans. Biomed. Eng.* 66(1207–1221), 2019. https://doi.org/10.1109/TBME.2018.2870542 (2019).
- Li, X., Guo, S., Shi, P., Jin, X. & Kawanishi, M. An endovascular catheterization robotic system using collaborative operation with magnetically controlled haptic force feedback. *Micromachines* 13, 505. https://doi.org/10.3390/mi13040505 (2022).
- Elayaperumal, S., Bae, J. H., Daniel, B. L. & Cutkosky, M. R. Detection of membrane puncture with haptic feedback using a tipforce sensing needle. *Rep. U. S.* 2014, 3975–3981. https://doi.org/10.1109/IROS.2014.6943121 (2014).
- Kim, U., Seok, D.-Y., Kim, Y. B., Lee, D.-H. & Choi, H. R. Development of a grasping force-feedback user interface for Surgical Robot System. In *IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS)*. https://doi.org/10.1109/iros.2016. 7759149 (2016).
- Zhang, L., Gu, S., Guo, S. & Tamiya, T. A magnetorheological fluids-based robot-assisted catheter/guidewire surgery system for endovascular catheterization. *Micromachines* 12, 640. https://doi.org/10.3390/mi12060640 (2021).
- Kitagawa, M., Dokko, D., Okamura, A. M. & Yuh, D. D. Effect of sensory substitution on suture-manipulation forces for robotic surgical systems. J. Thorac. Cardiovasc. Surg. 129, 151–158. https://doi.org/10.1016/j.jtcvs.2004.05.029 (2006).
- Mahvash, M. et al. Force-feedback surgical teleoperator: Controller design and palpation experiments. In 2008 Symposium on Haptic Interfaces for Virtual Environment and Teleoperator Systems 465–471. https://doi.org/10.1109/HAPTICS.2008.4479994 (2008)
- Talasaz, A., Trejos, A. L. & Patel, R. V. The role of direct and visual force feedback in suturing using a 7-DOF dual-arm teleoperated system. *IEEE Trans. Hapt.* 10, 276–287. https://doi.org/10.1109/TOH.2016.2616874 (2017).
- Howard, T. & Szewczyk, J. Assisting control of forces in laparoscopy using tactile and visual sensory substitution. In New Trends in Medical and Service Robots: Human Centered Analysis, Control and Design 151–164. https://doi.org/10.1007/978-3-319-30674-2_ 12 (Springer Cham, 2016).
- Juo, Y. Y. et al. Reducing retraction forces with tactile feedback during robotic total mesorectal excision in a porcine model. J. Robot. Surg. 16, 1083–1090. https://doi.org/10.1007/s11701-021-01338-w (2022).
- Binet, A. *et al.* Current perspectives in robot-assisted surgery. *Miner. Pediatr.* 70, 308–314. https://doi.org/10.23736/S0026-4946. 18.05113-7 (2018).
- Griffin, J. A., Zhu, W. & Nam, C. S. The role of haptic feedback in robotic-assisted retinal microsurgery systems: A systematic review. *IEEE Trans. Hapt.* 10, 94–105. https://doi.org/10.1109/TOH.2016.2598341 (2017).
- Gaudeni, C., Meli, L. & Prattichizzo, D. A novel pneumatic force sensor for robot-assisted surgery. In Haptics: Science, Technology, and Applications: 11th International Conference 587–599 (Springer International Publishing, 2018).
- Hagen, M. E., Meehan, J. J., Inan, I. & Morel, P. Visual clues act as a substitute for haptic feedback in robotic surgery. Surg. Endosc. 22, 1505–1508. https://doi.org/10.1007/s00464-007-9683-0 (2008).
- Amirabdollahian, F. et al. Prevalence of haptic feedback in robot-mediated surgery: A systematic review of literature. J. Robot. Surg. 12, 11–25. https://doi.org/10.1007/s11701-017-0763-4 (2018).
- El Rassi, I. & El Rassi, J. M. A review of haptic feedback in tele-operated robotic surgery. J. Med. Eng. Technol. 44, 247–254. https:// doi.org/10.1080/03091902.2020.1772391 (2020).
- McKechnie, T. et al. The Senhance surgical system in colorectal surgery: A systematic review. J. Robot. Surg. 2022, 1–10. https:// doi.org/10.1007/s11701-022-01455-0 (2022).
- Weber, B. M. & Eichberger, C. The benefits of haptic feedback in telesurgery and other teleoperation systems: A meta-analysis. In Universal Access in Human-Computer Interaction. Access to Learning, Health and Well-Being: 9th International Conference 394–405 (Springer International Publishing, 2015).
- Page, M. J. et al. The PRISMA 2020 statement: An updated guideline for reporting systematic reviews. Int. J. Surg. 88, 105906. https://doi.org/10.1016/j.ijsu.2021.105906 (2021).
- Kim, Y. et al. Virtual-reality cataract surgery simulator using haptic sensory substitution in continuous circular capsulorhexis. Annu. Int. Conf. IEEE Eng. Med. Bio. Soc. 2018, 1887–1890. https://doi.org/10.1109/EMBC.2018.8512708 (2018).
- Agarwal, R., Hussain, A., Skm, V. & Campolo, D. Haptic feedback system for postural adaptation during robotic rehabilitation of upper limb. *IEEE Int. Conf. Rehabil. Robot.* 2022, 1–6. https://doi.org/10.1109/ICORR55369.2022.9896531 (2022).
- 29. Cundy, T. P. *et al.* Force-Sensing Enhanced Simulation Environment (ForSense) for laparoscopic surgery training and assessment. *Surgery* **157**, 723–731. https://doi.org/10.1016/j.surg.2014.10.015 (2015).
- Van-Oosterhout, J. et al. Haptic assistance improves tele-manipulation with two asymmetric slaves. IEEE Trans. Hapt. 12, 141–153. https://doi.org/10.1109/TOH.2018.2873350 (2019).
- 31. Borenstein, M., Hedges, L., Higgins, J. & Rothstein, H. Introduction to Meta-Analysis 21-33 (Wiley, 2009).
- Wottawa, C. R. *et al.* Evaluating tactile feedback in robotic surgery for potential clinical application using an animal model. *Surg. Endosc.* 30, 3198–3209. https://doi.org/10.1007/s00464-015-4602-2 (2016).
- Medina, T., Sánchez-Meca, J., Marín-Martínez, F. & Botella, J. Assessing heterogeneity in meta-analysis: Q statistic or I² index?. Psych. Meth. 11, 193–206. https://doi.org/10.1037/1082-989X.11.2.193 (2006).
- Abiri, A. et al. Visual-perceptual mismatch in robotic surgery. Surg. Endosc. 31, 3271–3278. https://doi.org/10.1007/s00464-016-5358-z (2017).

- Abiri, A. et al. Artificial palpation in robotic surgery using haptic feedback. Surg. Endosc. 33, 1252–1259. https://doi.org/10.1007/ s00464-018-6405-8 (2019).
- Abiri, A. et al. Multi-modal haptic feedback for grip force reduction in robotic surgery. Sci. Rep. 9, 5016. https://doi.org/10.1038/ s41598-019-40821-1 (2019).
- Aggravi, M., Estima, D. A. L., Krupa, D. A. L. A., Misra, S. & Pacchierotti, C. Haptic teleoperation of flexible needles combining 3d ultrasound guidance and needle tip force feedback. *IEEE Robot. Autom. Lett.* 6, 4859–4866. https://doi.org/10.1109/LRA.2021. 3068635 (2021).
- Alleblas, C. C. J., Vleugels, M. P. H., Coppus, S. F. & Nieboer, T. E. The effects of laparoscopic graspers with enhanced haptic feedback on applied forces: A randomized comparison with conventional graspers. *Surg. Endosc.* 12, 5411–5417. https://doi.org/ 10.1007/s00464-017-5623-9 (2017).
- Bahar, L., Sharon, Y. & Nisky, I. Surgeon-centered analysis of robot-assisted needle driving under different force feedback conditions. Front. Neurorobot. 13, 85. https://doi.org/10.3389/fnbot.2019.00108 (2020).
- Bao, X., Guo, S., Xiao, N., Li, Y. & Shi, L. Compensatory force measurement and multimodal force feedback for remote-controlled vascular interventional robot. *Biomed. Microdev.* 20, 3. https://doi.org/10.1007/s10544-018-0318-0 (2019).
- Camara, M., Mayer, E., Darzi, A. & Pratt, P. Intraoperative ultrasound for improved 3D tumour reconstruction in robot-assisted surgery: An evaluation of feedback modalities. *Int. J. Med. Robot.* 15, 2. https://doi.org/10.1002/rcs.1973 (2019).
- Carobbio, A. L. C. et al. Transoral laser microsurgery: Feasibility of a new exoscopic HD-3D system coupled with free beam or fiber laser. Lasers Med. Sci. 36, 1865–1872. https://doi.org/10.1007/s10103-020-03221-w (2021).
- Chinello, F., Malvezzi, M., Prattichizzo, D. & Pacchierotti, C. A modular wearable finger interface for cutaneous and kinesthetic interaction: Control and evaluation. *IEEE Trans. Ind. Electron.* 67, 706–716. https://doi.org/10.1109/TIE.2019.2899551 (2020).
- Chowriappa, A. et al. A predictive model for haptic assistance in robot assisted trocar insertion. World Hapt. Conf. 2013, 121–126. https://doi.org/10.1109/WHC.2013.6548395 (2013).
- Currie, M. E. et al. The role of visual and direct force feedback in robotics-assisted mitral valve annuloplasty. Int. J. Med. Robot. 13, 3. https://doi.org/10.1002/rcs.1787 (2017).
- Dagnino, G. et al. Haptic feedback and dynamic active constraints for robot-assisted endovascular catheterization. U.S. Rep. 2018, 1770–1775. https://doi.org/10.1109/IROS.2018.8593628 (2018).
- 47. Dai, Y. *et al.* Biaxial sensing suture breakage warning system for robotic surgery. *Biomed. Microdev.* 21, 1. https://doi.org/10.1007/s10544-018-0357-6 (2019).
- Dalvand, M., Shirinzadeh, B., Nahavandi, S. & Smith, J. Effects of realistic force feedback in a robotic assisted minimally invasive surgery system. *Minim. Invasive Ther. Allied. Technol.* 23, 127–135. https://doi.org/10.3109/13645706.2013.867886 (2014).
- Ehrampoosh, S., Dave, M., Kia, M. A., Rablau, C. & Zadeh, M. H. Providing haptic feedback in robot-assisted minimally invasive surgery: A direct optical force-sensing solution for haptic rendering of deformable bodies. *Comput. Aided Surg.* 18, 129–141. https://doi.org/10.3109/10929088.2013.839744 (2013).
- Fichera, L., Pacchierotti, C., Olivieri, E., Prattichizzo, D. & Mattos, L. S. Kinesthetic and vibrotactile haptic feedback improves the performance of laser microsurgery. In 2016 IEEE Haptics Symp 59–64. https://doi.org/10.1109/HAPTICS.2016.7463156 (2016)
- Francone, A. The effect of haptic feedback on efficiency and safety during preretinal membrane peeling simulation. *Transl. Vis. Sci. Technol.* 8, 4. https://doi.org/10.1167/tvst.8.4.2 (2019).
- Gambaro, E. F., Zollo, L. & Guglielmelli, E. Development and experimental validation of a master interface with vibrotactile feedback for robotic telesurgery. In Proc. IEEE RAS EMBS Int. Conf. Biomed. Robot. Biomechatron 726–731. https://doi.org/10.1109/ BIOROB.2014.6913864 (2014).
- Gerena, E., Legendre, F., Vitry, Y., Régnier, S. & Haliyo, S. Improving optical micromanipulation with force-feedback bilateral coupling. *IEEE Int. Conf. Robot Autom.* 2020, 10292–10298 (2020).
- Gibo, T.L., Bastian, A. J. & Okamura, A. M. Grip force control during virtual object interaction: Effect of force feedback, accuracy demands, and training. *IEEE Trans. Hapt.* 7, 37–47. https://doi.org/10.1109/TOH.2013.60 (2014).
- Grammatikopoulou, M., Leibrandt, K. & Yang, G.-Z. Motor channelling for safe and effective dynamic constraints in minimally invasive surgery. Int. Conf. Intel. Robot. Syst. 4317–4323, 2016. https://doi.org/10.1109/IROS.2016.7759635 (2016).
- Jeong, S. & Tadano, K. Force feedback on hand rest function in master manipulator for robotic surgery. Int. Conf. Intel. Robot. Syst. 2021, 1815–1820 (2021).
- Jin, X. et al. Development of a tactile sensing robot-assisted system for vascular interventional surgery. IEEE Sens. J. 21, 12284– 12294. https://doi.org/10.1109/JSEN.2021.3066424 (2021).
- Karponis, D., Koya, Y., Miyazaki, R., Kanno, T. & Kawashima, K. Evaluation of a pneumatic surgical robot with dynamic force feedback. J. Robot. Surg. 13, 413–421. https://doi.org/10.1007/s11701-018-0878-2 (2019).
- Lim, S. C., Lee, H. K. & Park, J. Role of combined tactile and kinesthetic feedback in minimally invasive surgery. Int. J. Med. Robot. 11, 360–374. https://doi.org/10.1002/rcs.1625 (2015).
- Lopez, E., Zollo, L. & Guglielmelli, E. Teleoperated control based on virtual fixtures for a redundant surgical system. Int. Conf. Intel Robot. Syst. 2013, 450–455. https://doi.org/10.1109/IROS.2013.6696390 (2013).
- Mendelsohn, A. H. et al. Transoral robotic surgical proficiency via real-time tactile collision awareness system. Laryngoscope 130, 1-S17. https://doi.org/10.1002/lary.29034 (2020).
- Molinero, M. B. et al. Haptic guidance for robot-assisted endovascular procedures: Iimplementation and evaluation on surgical simulator. Int. Conf. Intel. Robot. Syst. 2019, 5398–5403. https://doi.org/10.1109/IROS40897.2019.8967712 (2019).
- Urias, M. G., Patel, N., Ebrahimi, A., Iordachita, I. & Gehlbach, P. L. Robotic retinal surgery impacts on scleral forces: In vivo study. *Transl. Vis. Sci. Technol.* 2020, 9. https://doi.org/10.1167/tvst.9.10.2 (2020).
- Nakazawa, A. *et al.* Feedback methods for collision avoidance using virtual fixtures for robotic neurosurgery in deep and narrow spaces. In *Proc. IEEE RAS EMBS Int. Conf. Biomed. Robot. Biomechatron* 247–252. https://doi.org/10.1109/BIOROB.2016.75236 32 (2016).
- Navkar, N. V., Deng, Z., Shah, D. J. & Tsekos, N. V. A framework for integrating real-time MRI with robot control: Application to simulated transapical cardiac interventions. *IEEE Trans. Biomed. Eng.* 60, 1023–1033. https://doi.org/10.1109/TBME.2012.22303 98 (2013).
- Olivieri, E. *et al.* Haptic feedback for control and active constraints in contactless laser surgery: Concept, implementation, and evaluation. *IEEE Trans. Hapt.* 11, 241–254. https://doi.org/10.1109/TOH.2017.2786243 (2018).
- Ouyang, Q. et al. Bio-inspired haptic feedback for artificial palpation in robotic surgery. *IEEE Trans. Biomed. Eng.* 68, 3184–3193. https://doi.org/10.1109/TBME.2021.3076094 (2021).
- Pacchierotti, C., Meli, L., Chinello, F., Malvezzi, M. & Prattichizzo, D. Cutaneous haptic feedback to ensure the stability of robotic teleoperation systems. *Int. J. Rob. Res.* 34, 1773–1787. https://doi.org/10.1177/0278364915603135 (2015).
- Pacchierotti, C., Prattichizzo, D. & Kuchenbecker, K. J. Cutaneous feedback of fingertip deformation and vibration for palpation in robotic surgery. *IEEE Trans. Biomed. Eng.* 63, 278–287. https://doi.org/10.1109/TBME.2015.2455932 (2016).
- Portolés, S. et al. Force control for tissue tensioning in precise robotic laser surgery. IEEE Int. Conf. Robot. Autom. 2015, 579–585. https://doi.org/10.1109/ICRA.2015.7139237 (2015).
- Power, M., Rafii-Tari, H., Bergeles, H., Vitiello, C. V. & Yang, G. Z. A cooperative control framework for haptic guidance of bimanual surgical tasks based on Learning From Demonstration. *IEEE Int. Conf. Robot. Autom.* 2015, 5330–5337. https://doi.org/10.1109/ ICRA.2015.7139943 (2015).

14

- Quek, Z. F., Provancher, W. R. & Okamura, A. M. Evaluation of skin deformation tactile feedback for teleoperated surgical tasks. *IEEE Trans. Hapt.* 12, 102–113. https://doi.org/10.1109/TOH.2018.2873398 (2019).
- Saracino, A. *et al.* Haptic feedback in the da Vinci Research Kit (dVRK): A user study based on grasping, palpation, and incision tasks. *Int. J. Med. Robot.* 2019, 15. https://doi.org/10.1002/rcs.1999 (2019).
- Saracino, A., Oude-Vrielink, T. J. C., Menciassi, A., Sinibald, I. E. & Mylonas, G. P. Haptic intracorporeal palpation using a cabledriven parallel robot: A user study. *IEEE Trans. Biomed. Eng.* 67, 3452–3463. https://doi.org/10.1109/TBME.2020.2987646 (2020).
- Seung, S. et al. Virtual wall-based haptic-guided teleoperated surgical robotic system for single-port brain tumor removal surgery. Proc. Inst. Mech. Eng. H. 231, 3–19. https://doi.org/10.1177/0954411916676218 (2017).
- Song, Y., Guo, S., Zhang, L. & Yu, M. Haptic feedback in robot-assisted endovascular catheterization. Int. Conf. Mechatron. Autom. 2017, 404–409. https://doi.org/10.1109/ICMA.2017.8015851 (2017).
- Tahir, A., Iqbal, H., Usman, M., Ghaffar, A. & Hafeez, A. Cardiac X-ray image-based haptic guidance for robot-assisted coronary intervention: A feasibility study. Int. J. Comput. Assist. Radiol. Surg. 17, 531–539. https://doi.org/10.1007/s11548-022-02563-w (2022).
- Yin, X., Guo, S. & Song, Y. Magnetorheological fluids actuated haptic-based teleoperated catheter operating system. *Micromach. Basel* 3, 859. https://doi.org/10.3390/mi9090465 (2018).
- 79. Wirtz, M. A. Fail-Safe-N-Methode. https://dorsch.hogrefe.com/stichwort/fail-safe-n-methode (2022).
- Nitsch, V. & Färber, B. A meta-analysis of the effects of haptic interfaces on task performance with teleoperation systems. *IEEE Trans. Hapt.* 6, 387-398. https://doi.org/10.1109/TOH.2012.62 (2013).
- 81. Weber, B. & Schneider, S. The effects of force feedback on surgical task performance: A meta-analytical integration in Haptics. In Neuroscience, Devices, Modeling, and Applications: 9th International Conference 150–157 (Springer, 2014).
- Meccariello, G. et al. An experimental study about haptic feedback in robotic surgery: May visual feedback substitute tactile feedback?. J. Robot. Surg. 10, 57-61. https://doi.org/10.1007/s11701-015-0541-0 (2016).
- Okumura, Y. et al. A systematical analysis of in vivo contact forces on virtual catheter tip/tissue surface contact during cardiac mapping and intervention. J. Cardiovasc. Electrophysiol. 19, 632–640. https://doi.org/10.1111/j.1540-8167.2008.01135.x (2008).

Author contributions

Study concept and design: M.B., M.F., B.M.W. Acquisition of data: M.B. Analysis and interpretation of data: M.B. Drafting of the manuscript: M.B. Critical revision of the manuscript: M.F., B.M.W. Statistical analysis: M.B. Administrative, technical, or material support: M.F., B.M.W. Study supervision: M.F., B.M.W.

Funding

Open Access funding enabled and organized by Projekt DEAL.

Competing interests

The authors declare no competing interests.

Additional information

Supplementary Information The online version contains supplementary material available at https://doi.org/10.1038/s41598-023-46641-8.

Correspondence and requests for materials should be addressed to M.F.

Reprints and permissions information is available at www.nature.com/reprints.

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit http://creativecommons.org/licenses/by/4.0/.

© The Author(s) 2023

Terms and Conditions

Springer Nature journal content, brought to you courtesy of Springer Nature Customer Service Center GmbH ("Springer Nature").

Springer Nature supports a reasonable amount of sharing of research papers by authors, subscribers and authorised users ("Users"), for smallscale personal, non-commercial use provided that all copyright, trade and service marks and other proprietary notices are maintained. By accessing, sharing, receiving or otherwise using the Springer Nature journal content you agree to these terms of use ("Terms"). For these purposes, Springer Nature considers academic use (by researchers and students) to be non-commercial.

These Terms are supplementary and will apply in addition to any applicable website terms and conditions, a relevant site licence or a personal subscription. These Terms will prevail over any conflict or ambiguity with regards to the relevant terms, a site licence or a personal subscription (to the extent of the conflict or ambiguity only). For Creative Commons-licensed articles, the terms of the Creative Commons license used will apply.

We collect and use personal data to provide access to the Springer Nature journal content. We may also use these personal data internally within ResearchGate and Springer Nature and as agreed share it, in an anonymised way, for purposes of tracking, analysis and reporting. We will not otherwise disclose your personal data outside the ResearchGate or the Springer Nature group of companies unless we have your permission as detailed in the Privacy Policy.

While Users may use the Springer Nature journal content for small scale, personal non-commercial use, it is important to note that Users may not:

- 1. use such content for the purpose of providing other users with access on a regular or large scale basis or as a means to circumvent access control;
- 2. use such content where to do so would be considered a criminal or statutory offence in any jurisdiction, or gives rise to civil liability, or is otherwise unlawful;
- 3. falsely or misleadingly imply or suggest endorsement, approval, sponsorship, or association unless explicitly agreed to by Springer Nature in writing;
- 4. use bots or other automated methods to access the content or redirect messages
- 5. override any security feature or exclusionary protocol; or
- 6. share the content in order to create substitute for Springer Nature products or services or a systematic database of Springer Nature journal content.

In line with the restriction against commercial use, Springer Nature does not permit the creation of a product or service that creates revenue, royalties, rent or income from our content or its inclusion as part of a paid for service or for other commercial gain. Springer Nature journal content cannot be used for inter-library loans and librarians may not upload Springer Nature journal content on a large scale into their, or any other, institutional repository.

These terms of use are reviewed regularly and may be amended at any time. Springer Nature is not obligated to publish any information or content on this website and may remove it or features or functionality at our sole discretion, at any time with or without notice. Springer Nature may revoke this licence to you at any time and remove access to any copies of the Springer Nature journal content which have been saved.

To the fullest extent permitted by law, Springer Nature makes no warranties, representations or guarantees to Users, either express or implied with respect to the Springer nature journal content and all parties disclaim and waive any implied warranties or warranties imposed by law, including merchantability or fitness for any particular purpose.

Please note that these rights do not automatically extend to content, data or other material published by Springer Nature that may be licensed from third parties.

If you would like to use or distribute our Springer Nature journal content to a wider audience or on a regular basis or in any other manner not expressly permitted by these Terms, please contact Springer Nature at

onlineservice@springernature.com