

Evaluating subsampling strategies for sEMG-based prediction of voluntary muscle contractions

Risto Kõiva
Neuroinformatics Group,
Bielefeld University
D-33615 Bielefeld, Germany
email: risto.koiva@uni-bielefeld.de

Barbara Hilsenbeck and Claudio Castellini
Robotics and Mechatronics Center
DLR — German Aerospace Center
D-82234 Weßling, Germany
email: claudio.castellini@dlr.de

Abstract—In previous work we showed that some human Voluntary Muscle Contractions (VMCs) of high interest to the prosthetics community, namely finger flexions/extensions and thumb rotation, can be effectively predicted using muscle activation signals coming from surface electromyography (sEMG). In this paper we study the effectiveness of various subsampling strategies to limit the size of the training data set, with the aim of extending the approach to an online VMC-prediction system whose main application will be force-controlled hand prostheses. We performed an experiment in which 10 able-bodied participants flexed and extended their fingers according to a visual stimulus, while muscle activations and VMCs (represented as synergistic fingertip forces) were gathered using sEMG electrodes and a custom-built measurement device. A Support Vector Machine (SVM) was trained on a fixed-sized subset of the collected data, obtained using seven different subsampling strategies. The SVM was then tested on subsequent new data. Our experimental results show that two subsampling strategies attain a prediction error as low as 6% to 12%, which is comparable to the error values obtained in our previous work when the entire data set was used and processed offline.

I. INTRODUCTION

Over the past 30 years surface electromyography (sEMG) has received increased interest from the rehabilitation research community as an input modality for controlling advanced hand prostheses (see [1], and also the surveys in [2], [3]). Traditionally two sEMG electrodes were used to capture muscular signals that triggered movements of a simple one degree-of-freedom open/close type gripper when predefined thresholds were exceeded. However, with the appearance of polyarticulated prostheses such as, e.g., Otto Bock's Michelangelo¹ and RSL Steeper's BeBionic² hands, more sophisticated control solutions to actuate the single fingers are being investigated. In particular, sEMG patterns are classified according to the desired hand posture, leading to a more natural form of open-loop control. This approach has, in turn, at least three disadvantages: (a) it forces the user to select from a limited set of predefined grip types; (b) the stability of the prosthesis is starkly affected by the accuracy of the classifier; and (c) it enforces no control over the *force* the prosthesis applies.

In order to circumvent these problems, we have recently proposed [4] a radical shift: to use *Voluntary Muscle Con-*

tractions (VMCs), rather than hand position configurations, as the target values of machine-learning-based prediction. VMCs are defined as voluntary actions of muscle contraction which produce measurable effects, in our case forces. Typically, when a participant is asked to attain a specific finger force, for example to voluntarily flex the index finger to 50% of its maximum force, adjacent fingers will also change posture and produce side-effect forces [5]. Using VMCs allows us to isolate the finger of interest and associate synergistic muscle activations with it.

VMCs can be quantitatively elicited from any participant using visual stimuli, which is a standard experimental paradigm for this kind of investigation. Moreover, a set of VMCs of interest can be selected according to the degrees of freedom afforded by a particular prosthesis, therefore simplifying the control task. Finally, VMCs provide a continuous representation of the action space, independent of the participant's anatomy (the muscles and/or motor-units involved). This means that their prediction enables force-based control over an infinite manifold of configurations. We argue that this will maximize the dexterity of control and in the future allow prosthetic wearing amputees to perform a wider variety of activities in their daily lives through enhanced control of prosthesis hardware.

In this paper we build on the results obtained in our previous work [4], extending the prediction to unseen data using a training set that is strictly limited in size. In general cases, in which the prediction model is trained online, a strategy to limit the amount of training samples is normally required to reduce the computational effort. Such models typically obtain much worse results on unseen data, largely due to the well-known issue of the sEMG signal drifting and changing over time [6], [7]. Online processing of sEMG signals for classification has yielded interesting but sub-optimal results in recent times, for instance in [8] where an average classification accuracy of 79% on amputees was reported. This is far below the results obtained for able-bodied participants in the offline setting.

In the experiment ten able-bodied participants were instructed to apply finger-force patterns associated with VMCs of interest during a data-gathering phase, while wearing sEMG electrodes around the forearm to capture muscle activations. The obtained data sets were subsampled using a number of

¹www.ottobock.com

²www.bebionic.com

schemes and six Support Vector Machines (SVM) [9], [10] were trained on the reduced data set. In the final phase, VMCs were predicted (as synergistic fingertip forces) using the obtained models. The best results are obtained by ignoring samples which are too close (in force space) to previous ones. The resulting prediction errors range between 6.39% and 11.54% of the force ranges; these values are in line with those we previously obtained in a similar, but offline, experiment in [4]. We note that none of the strategies we employed emerged as a clear winner, suggesting that there exists some flexibility in choosing a subsampling strategy.

In the next section we describe the experimental setup and protocol in detail. In Sec. III we discuss the analysis methods and tested subsampling strategies. Sec. IV presents the experimental results, and finally in Sec. V we draw conclusions and outline future work.

II. EXPERIMENT DESCRIPTION

The goal of the experiment was to predict VMCs as synergistic combinations of fingertip forces. To this aim the participants applied finger forces and their forearm muscle activations and the forces produced by the fingertips were recorded. The setup consisted of three main parts: the sEMG electrodes, the finger force capturing sensor rig and visual stimuli presented on a monitor [Fig. 1].

A. Surface electromyography

Ten OttoBock *MyoBock 13E200* sEMG electrodes were used to capture the muscle activations on the participant's forearm. These electrodes provide an amplified, bandpass filtered and rectified signal. They were strapped around the participant's forearm, just below the elbow, using an adjustable hook-and-loop band. Five electrodes were placed on the dorsal side, and the other five were placed on the ventral side of the forearm. All electrodes were evenly distributed without any respect to intrinsic muscles (the region of the ulna was avoided since no strong muscular signal is available there).

B. Fingertip forces

In general, a single VMC generates a simultaneous change of flexion/extension force in several fingertips, making an accurate synchronous gathering of forces from all fingers a necessity. To this end we use our custom built Finger-Force Linear Sensor (FFLS) [11]. The FFLS measures single finger flexion and extension forces³ of the index, middle, ring and little fingers, using an individually calibrated $\pm 100\text{N}$ industrial strain gauge based force sensor for each finger. Thumb rotation and adduction/abduction forces are captured using a single radial dual-axis calibrated strain gauge sensor with a similar $\pm 100\text{N}$ range⁴. These sensors are characterized by their high signal repeatability, minimal drift over time and almost perfect

³Finger flexion and extension refer to pushing the finger down and pulling the finger up respectively.

⁴Thumb rotation is an intuitive description of the thumb rotational motion, and thumb adduction/abduction refers to drawing the thumb closer or pushing the thumb further away from the palm.




Fig. 1. The experimental setup. Visual stimuli, consisting of an array of animated, colored bars (upper part of the picture) is shown to a participant. The participant's fingertips are attached to the finger force capturing device (Finger-Force Linear Sensor (FFLS) [11]). The sEMG electrodes are clearly visible on the forearm just below the elbow.

linearity and non-existent hysteresis with both parameters guaranteed to deviate not more than 0.3%. Each strain gauge force sensor is connected to a dedicated industrial strain gauge amplifier, which converts the applied force into an analog voltage.

All sensors are adjustable for various hand sizes and were mounted rigidly after positional adjustments. The fingers are tightened to the sensors using hook-and-loop bands with minimal slack. Custom made gypsum casts are used for backlash-




Fig. 2. The stimulus and typical data collected during the training phase. **(top panel)** The sequence of the stimuli, six single-finger VMCs and three multi-finger VMCs. **(center panel)** The corresponding force data recorded by the FFLS — colors denote the force sensors and the force values are mostly negative since we defined flexion as the negative direction of the sensor axes. The six single- and three multi-finger VMCs are clearly seen; for instance, flexion of the index appears roughly between 140s and 180s as a series of five blue negative force peaks; the pinch grip (combination of index flexion and thumb adduction) appears as a series of combined blue and yellow peaks between 290s and 330s. **(bottom panel)** Two typical sEMG signals (red denoting an electrode placed on the dorsal side of the forearm, blue on the ventral side) corresponding to the elicited VMCs.

free (to ensure data is recorded for all movements) attachment of the participant’s thumbs to the dual-axis sensor. Both the sEMG electrodes and the FFLS are connected to a 12-bit resolution multichannel digital acquisition card that gathers data with a sampling frequency of 25Hz. This relatively low sampling rate was sufficient because the stimulated individual actions lasted for 5 to 10 seconds each and the sEMG electrodes provided a filtered signal.

C. Visual stimulus

A graphical user interface was displayed on a large monitor in front of the participant. The application was programmed to be used intuitively without prior training, and makes extensive use of wizards that led the participants’ through the different phases of the experiment, namely calibration, data gathering and prediction. The requested VMC was graphically presented using six red colored bars, one for each considered finger movement direction (flexion of the index, middle, ring and little fingers, plus thumb rotation and adduction). The height of the bars denotes the intensity of the targeted finger force. Next to the stimulus, blue colored bars show the measured fingertip forces and thus provide the participant with real-time visual feedback. During the prediction phase, the predicted finger forces are shown in an additional, third bar, colored in green.

D. Experimental protocol

The experiment consists of the calibration, data gathering and prediction phases. In the calibration phase, the idle and the

maximum forces of each finger are recorded and averaged over a period of two seconds each. During the phases following the calibration, the requested force is limited to a fixed fraction of the captured maximum force. The top panel of Fig. 2 shows the nine requested VMCs during the data gathering phase. Six single-finger and three multi-finger VMCs were requested, the latter ones representing common grasp postures used in everyday life and resembling common grip patterns [12]. We concentrated on a pinch grip, an index finger pointing gesture and a power grasp. With the pinch grip (simultaneous thumb adduction and index flexion) small objects like a pen can be picked up. The index pointing gesture (flexion of the middle, ring and little finger) is used for pressing buttons and switches. Big objects, for example a bottle, are typically held using a power grasp, which is performed by flexing all fingers simultaneously.

During the data gathering phase, the stimulus increased from 0 to 80% of the participant’s maximum achievable force from the calibration phase for the single-finger VMCs and from 0 to 60% for the multi-finger VMCs, and then decreased back to 0 in a total of six seconds. Two seconds of rest were provided between VMCs in order to avoid muscle fatigue. Each VMC was repeated five times, which resulted in a total of 45 VMCs and a data gathering phase duration of ≈ 7 minutes [Fig. 2]. The data gathering phase was followed by a prediction phase, during which all VMCs were repeated two times (total duration, 2.8 minutes). The total experiment duration lasted ≈ 15 minutes, including the positional adjustment of all

sensors, the calibration, data gathering and prediction phases.

The system was tested with 10 able-bodied participants (all right-handed; 9 male, 1 female; aged 27.40 ± 4.41 years, min: 23, max: 29). The participants sat on an office chair, which was adjusted for maximum comfort, but ensuring that the armrest was kept at the same height as the FFLS. By slightly reclining the chair, we made it difficult for participants to use their body weight to generate extra force. They were also explicitly instructed not to try this. No fatigue was reported for the single-finger VMCs. However, it is interesting to note that all participants reported that it was more difficult to follow the stimulus for the multi-finger VMCs, even though the required forces were reduced to only 60% of the maximum single finger force value. Indeed, some participants did not even manage to reach the requested 60% force level simultaneously with multiple fingers. Nevertheless, it is important to note that ground truth is always available in terms of fingertip forces synchronized to the sEMG, therefore avoiding the need for the participant to exactly match the stimuli (which would be essentially impossible).

III. DATA ANALYSIS

The maps from sEMG to the six forces representing the VMCs were built using six Support Vector Machines (SVM, [9], [13]) with Radial-Basis-Function kernels in ϵ -regression mode [14]. The sEMG signals, one for each electrode, define the \mathbb{R}^{10} input space, while the six force values representing the six VMCs define the \mathbb{R}^6 output space. The data collected during the data gathering phase was pre-processed, subsampled and then used to train the six SVMs. Subsequently, the SVMs were used to predict in real-time (at 25Hz) the force applied by the participants during the prediction phase. The optimal SVM hyperparameters C and γ were determined through cross-validation and grid-search of one recorded dataset in an initial round of experiments (optimal values: $C = 10, \gamma = 0.1$; ϵ was set at one tenth of the force signal's standard deviation). The training data was normalized by subtracting the means and dividing by the standard deviations. Normalization of the testing set during the prediction phase was performed in the same way using the statistics of the training set.

A. Preprocessing

The force data was preprocessed using a 1st order Butterworth low-pass filter with a 2Hz cutoff frequency, which is similar to our previous work [4]. For the sEMG data, a cutoff frequency of 1Hz was chosen. The selected filter frequencies were determined through visual inspection of the spectral contents to ensure that all important information was retained. With a sampling rate of 25Hz, this resulted in $\approx 10,500$ sEMG/force pairs for the training phase.

B. Subsampling Strategies

Several subsampling strategies were employed to build training data sets out of the data collected during the data gathering phase. Subsampling ensures that the computational requirements needed by an online system are not exceeded,

i.e., it keeps the training set at a fixed size even though in an online setting the input data set could be extremely large. The main idea behind the subsampling strategies we compared is that of *distance between pairs of samples*, a choice justified as follows: a map from sEMG to forces represents a relationship between two physiological events (muscle activation and the force exerted at the fingertips); therefore, it is reasonable to assume, according to the literature [15] and using our intuition, that the relationship is a smooth one. According to this argument, samples which are too close to previously recorded ones can be neglected. For our approach, closeness is represented as an inter-sample distance in either force or sEMG space.

More formally, let S be the set of sEMG/force samples collected during the data gathering phase. A subset $S' \subset S$ was built according to one of the following subsampling strategies:

- 1) samples were randomly picked from S (*random subsampling*);
- 2) samples were randomly picked from S , but this time they were only selected if the forces were above 75% of the sampled maximum force during the data gathering phase (*random subsampling with preselection*);
- 3) samples picked from S such that $\|s_i - s_j\| > d, s_i, s_j \in S'$, for any $i, j = 1, \dots, |S|$, where d is a threshold that represents the minimum inter-sample distance allowed (*Euclidean distance subsampling of sEMG data*);
- 4) same as above, but the distance test is done in the output space on the force data (*Euclidean distance subsampling of force data*);
- 5) same as item 3 but the Mahalanobis distance is used instead of the Euclidean distance, i.e., given two samples s_1, s_2 , $MD(s_1, s_2) = \sqrt{(s_1 - s_2)^T C^{-1} (s_1 - s_2)}$ where C^{-1} is the inverse covariance matrix of S (*Mahalanobis distance subsampling of sEMG data*);
- 6) same as item 4 but the Mahalanobis distance is used (*Mahalanobis distance subsampling of force data*);
- 7) sEMG samples which yield no hash collision when entered into a standard hash table of predefined size. Hashing was done using 20 equidistal bins in sEMG space (*hashing*).

Of course for each strategy one needs to decide the size of the resulting set, $|S'|$; in the case of random subsampling (strategies 1 and 2 above), this can be directly chosen, while in the other cases d must be chosen each time in order for $|S'|$ to be close to the desired size. In both cases, in order to check the effectiveness of each strategy, we let $|S'|$ range over several values between 300 and 1500 (this upper value was limited by the CPU speed of the testing machine), in each case checking the attained error rate on the set $S - S'$ [Fig. 3]. As an error measure, we chose the Root-Mean-Square error, normalized over the range of the observed force values (NRMSE).

The best results were achieved using *random subsampling with preselection* and *Mahalanobis distance subsampling of force data* (strategies 2 and 6). Fig. 4 shows the detailed




Fig. 3. Error rates for each subsampling strategy as $|S'|$ increases.

comparison of these two strategies. For *random subsampling with preselection*, the test was repeated 50 times and the error was averaged (in the figure, one standard deviation is shown). The comparison of these strategies reveals no clear winner for the three VMCs shown, and this was also the case for the remaining VMCs (not depicted). In order to build the sEMG-to-VMC maps for the prediction phase of our experiment, we chose *Mahalanobis distance subsampling of force data*, mainly due to its non-random nature; given the same S , this strategy will always yield the same S' .

IV. RESULTS

NRMSE was used as a measure to evaluate the performance of the models obtained from S' on the data collected during the prediction phase. Table I shows the results for all participants and all tested finger movements. This means that the statistics shown for each finger movement (e.g., Index) include both single and multiple VMCs.

The best accuracy was obtained by Participant #7, with error values ranging between 3.03% and 9.76%, and the worst accuracy was obtained by Participant #8, with error values ranging between 8.23% and 16.27%. The finger movement that was best predicted was flexion of the ring finger, with




Fig. 4. Error rates of the two best strategies: random subsampling with preselection and Mahalanobis distance subsampling of force data, for three typical VMCs (flexion of the index and little fingers and thumb adduction).

error values from 3.03% to 11.79%, whereas the thumb adduction prediction performed worst, with values from 4.60% to 16.27%. The fact that the thumb adduction was the hardest finger movement to predict is not surprising, given that muscles devoted to its movement lie mainly within the hand and are far from the sEMG gathering sensors. It is on the other hand comforting to see that the thumb rotation, an extremely important VMC, can be predicted with similar accuracy to the other finger VMCs ($8.90\% \pm 0.93\%$).

Fig. 5 displays a typical comparison between measured and predicted force values for the data from Participant #3. As can be seen, the prediction errors for single- and multi-finger VMCs are comparable. In some cases the prediction for the multi-finger VMCs (the peaks starting from 120 second mark) is even slightly better than for the single finger movements. This is probably due to the fact that it is impossible to simultaneously apply equal force with many fingers [5], which results in a large Mahalanobis distance through variances in the force patterns. Therefore, during the training phase, a higher percentage of data associated with multiple VMCs could be favored for selection into S' to train the SVM models.

V. CONCLUSIONS AND FUTURE WORK

The existence of a stable relationship between sEMG and some VMCs of interest, namely finger flexions and extensions and thumb rotation, has already been shown in [4]. The VMCs were chosen to match the degrees of freedom of modern, polyarticulated hand prostheses. In this paper, we

TABLE I
NORMALIZED ROOT-MEAN-SQUARE ERROR (NRMSE) FOR ALL PARTICIPANTS AND FINGERS TESTED, PLUS RELATED STATISTICS.

Participant#	NRMSE [%]						mean \pm std.
	Index	Middle	Ring	Little	Th.rot.	Th.add.	
1	9.94	9.22	9.14	6.56	10.15	15.91	10.15 \pm 3.10
2	6.95	9.27	8.72	9.18	8.72	4.97	7.97 \pm 1.69
3	9.29	10.80	7.58	4.74	7.36	4.60	7.40 \pm 2.45
4	11.16	7.04	6.51	7.41	8.43	11.86	8.74 \pm 2.25
5	9.40	8.80	5.20	6.22	9.06	14.71	8.90 \pm 3.31
6	10.32	7.79	6.05	6.19	8.59	13.87	8.80 \pm 2.95
7	9.76	5.61	3.03	3.96	9.23	6.73	6.39 \pm 2.73
8	13.07	11.21	8.23	9.79	10.74	16.27	11.54 \pm 2.81
9	11.90	12.08	5.15	7.39	8.74	6.03	8.55 \pm 2.93
10	7.77	9.70	11.79	6.42	8.01	6.25	8.32 \pm 2.11
mean \pm std.	9.96 \pm 1.72	9.15 \pm 1.87	7.14 \pm 2.36	6.79 \pm 1.69	8.90 \pm 0.93	10.12 \pm 4.58	




Fig. 5. Recorded and predicted force values of all considered VMCs during one representative trial. All VMCs performed in the training phase are repeated two more times for testing. The data from 0 to 120 seconds refers to force values gathered during single-finger VMCs phases, and the data after 120 seconds refers to the force values captured while participants' were performing multi-finger VMCs.

extend our previous work towards having online training of SVM models using limited-sized training datasets. In the described experiment, 10 able-bodied participants were asked to apply graded VMCs according to a visual stimulus. Several subsampling strategies were applied to the obtained datasets, and models obtained from these subsets were used to predict the same VMCs obtained during a further session. Our results demonstrate that subsampling strategies can be used to obtain error rates similar to those obtained on the entire dataset [4] with values ranging from 6% to 12%. Whether this accuracy is usable in practice is still the subject of future research, but our approach is promising.

A number of factors had to be considered. Firstly, the usage of a strategy for subsampling is just one (and probably the simplest) of the possible ways to bound a training set of a machine learning method in an online setting. New data is always available, and no method can afford to work on an ever-growing dataset. This is the main motivation behind this research. Secondly, the selection of a particular subsampling strategy proved to be not as critical as we had expected. This finding needs further investigation; for example, it is

not entirely clear which strategy scales up to much larger datasets, collected over hours and days. Random strategies, on the other hand, would almost certainly prove sensitive to the frequency content of the stimulus. Nonetheless, our approach is promising, revealing that online training and prediction with a limited training dataset is viable.

As far as timing is concerned, any method supposed to work online needs to enforce fast prediction, at least matching the capability of the target control system (i.e., a prosthesis) or cases in which a simulated model is given as visual output to a participant, should achieve an update rate in excess of ≈ 25 Hz. In our case, the prediction takes roughly $100\mu\text{s}$ with 2000 support vectors on a current typical desktop computer, which results in more than an adequate performance. We note however that the *training* phase, even with the reduced dataset, may take up to 10s.

A practical application of this system should take into account the limited capabilities of amputees to apply graded forces with the residual muscles of the missing limb. To this aim we are studying the possibility of training on only null and maximum forces and without using any force sensor; initial

results reveal that this strategy is promising. Generalization of single-finger VMC to multi-finger ones is still an open issue and is being investigated as well.

We note that our choice of the NRMSE as a measure of error might not be optimal and that some alternative form of locally-weighted error (see, e.g., [15]) could be more suited to the problem. However, the final benchmark should be a physiological one, such as, e.g., the success rate of human participants engaged in a box-and-block task [16] while using a prosthetic device controlled by various algorithms. Therefore, work is underway to test our approach using a 3D hand model in order to check whether amputees can perform the exercise with the same level of performance as was achieved by able-bodied participants during this experiment. Based on previous results, e.g., [17], we are confident of success. Finally, there are plans to test the stability of our system with a robotic hand.

ACKNOWLEDGMENT

This work was partially supported by the Swiss National Science Foundation Sinergia project #132700 NINAPRO (Non-Invasive Adaptive Hand Prosthetics), by the European Commission project The Hand Embodied (FP7-248587) and by the DFG Center of Excellence EXC 277: Cognitive Interaction Technology (CITEC). We would like to thank Patrick van der Smagt of the Technische Universität München and Michael Suppa of the DLR for their continual financial, moral and logistic support. We thank Mr. Johann Buchner of the DLR, for providing us with some electro-technical parts for the experiments. Finally, we thank Jonathan Maycock for his input and proofreading.

REFERENCES

- [1] G. N. Saridis and T. P. Gootee, "EMG pattern analysis and classification for a prosthetic arm," *IEEE Transactions on Biomedical Engineering*, vol. BME-29, no. 6, pp. 403–412, June 1982.
- [2] S. Micera, J. Carpaneto, and S. Raspopovic, "Control of hand prostheses using peripheral information," *IEEE Reviews in Biomedical Engineering*, vol. 3, pp. 48–68, 2010.
- [3] B. Peerdeman, D. Boere, H. Witteveen, R. Huis in 't Veld, H. Hermens, S. Stramigioli, H. Rietman, P. Veltink, and S. Misra, "Myoelectric forearm prostheses: State of the art from a user-centered perspective," *Journal of Rehabilitation Research & Development*, vol. 48, no. 6, pp. 719–738, 2011.
- [4] C. Castellini and R. Kõiva, "Using Surface Electromyography to Predict Single Finger Forces," in *IEEE RAS & EMBS International Conference on Biomedical Robotics and Biomechanics (BioRob 2012)*, Rome, Italy, June 2012.
- [5] V. M. Zatsiorsky, Z.-M. Li, and M. L. Latash, "Enslaving effects in multi-finger force production," *Experimental Brain Research*, vol. 131, pp. 187–195, 2000. [Online]. Available: <http://dx.doi.org/10.1007/s002219900261>
- [6] R. Merletti, M. Avenaggiato, A. Botter, A. Holobar, H. Marateb, and T. Vieira, "Advances in surface EMG: Recent progress in detection and processing techniques," *Critical reviews in biomedical engineering*, vol. 38, no. 4, pp. 305–345, 2011.
- [7] R. Merletti, A. Botter, C. Cescon, M. Minetto, and T. Vieira, "Advances in surface EMG: Recent progress in clinical research applications," *Critical reviews in biomedical engineering*, vol. 38, no. 4, pp. 347–379, 2011.
- [8] C. Cipriani, C. Antfolk, M. Controzzi, G. Lundborg, B. Rosén, M. Carrozza, and F. Sebelius, "Online myoelectric control of a dexterous hand prosthesis by transradial amputees," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 19, no. 3, pp. 260–270, June 2011.
- [9] B. E. Boser, I. M. Guyon, and V. N. Vapnik, "A training algorithm for optimal margin classifiers," in *Proceedings of the 5th Annual ACM Workshop on Computational Learning Theory*, D. Haussler, Ed. ACM press, 1992, pp. 144–152.
- [10] N. Cristianini and J. Shawe-Taylor, *An Introduction to Support Vector Machines (and Other Kernel-Based Learning Methods)*. CUP, 2000.
- [11] R. Kõiva, B. Hilsenbeck, and C. Castellini, "FFLS: An accurate linear device for measuring synergistic finger contractions," in *IEEE Engineering in Medicine & Biology Society (EMBC 2012)*, San Diego, USA, Aug. 2012.
- [12] M. Cutkosky, "On grasp choice, grasp models, and the design of hands for manufacturing tasks," *Robotics and Automation, IEEE Transactions on*, vol. 5, no. 3, pp. 269–279, June 1989.
- [13] V. N. Vapnik, *Statistical Learning Theory*. New York: John Wiley and Sons, 1998.
- [14] A. J. Smola and B. Schölkopf, "A tutorial on support vector regression," *Statistics and Computing*, vol. 14, pp. 199–222, 2004.
- [15] C. Castellini and P. van der Smagt, "Surface EMG in advanced hand prosthetics," *Biological Cybernetics*, vol. 100, no. 1, pp. 35–47, 2009.
- [16] V. Mathiowetz, G. Volland, N. Kashman, and K. Weber, "Adult norms for the box and block test of manual dexterity," *American Journal of Occupational Therapy*, vol. 39, no. 6, pp. 386–391, 1985.
- [17] C. Castellini, E. Gruppioni, A. Davalli, and G. Sandini, "Fine detection of grasp force and posture by amputees via surface electromyography," *Journal of Physiology (Paris)*, vol. 103, no. 3-5, pp. 255–262, 2009.